

Please enter these **calibration parameters** and the **Lot No.** into the BioLection software!

pH calibration parameters Lot No. 1704 (BioLector® II/Pro, filter module ID-202)

Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
ϕ min	65.88	65.77	65.66	65.55	65.44	65.33	65.22
ϕ max	20.12	20.10	20.08	20.06	20.05	20.03	20.01
dpH	0.49	0.49	0.49	0.49	0.49	0.49	0.49
pH ₀	6.75	6.75	6.74	6.73	6.73	6.72	6.72
Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
ϕ min	65.10	64.99	64.88	64.77	64.66	64.55	64.44
ϕ max	19.99	19.97	19.95	19.93	19.92	19.90	19.88
dpH	0.49	0.49	0.49	0.49	0.49	0.49	0.49
pH ₀	6.71	6.70	6.70	6.69	6.68	6.68	6.67
Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
ϕ min	64.33	64.22	64.10	63.99	63.88	63.77	63.66
ϕ max	19.86	19.84	19.82	19.80	19.79	19.77	19.75
dpH	0.49	0.49	0.49	0.49	0.49	0.49	0.49
pH ₀	6.67	6.66	6.65	6.65	6.64	6.64	6.63

pH sensor properties

Dynamic range	pH 4.05 - 8.80
Resolution	Up to 0.01 pH (software)
Accuracy	± 0.25 pH at pH 4.55 - 5.20; ± 0.1 pH at pH 5.20 – 7.60; ± 0.25 pH at pH 7.60 - 8.30 (batch calibration)
Response time (t90)	At 25 °C < 30 s
Drift at pH = 7	< 0.005 pH per day (sampling interval of 6 min)
Temperature range	5 °C to 50 °C
Compatibility	Aqueous solutions, ethanol, methanol (max. 5 % v/v)
Sensor stability	sensor material can be degraded by some microorganisms
Cross-sensitivity	Reduced to ionic strength (salinity); high concentration of fluorescent molecules in the visible range can interfere (GFP, (e)YFP); complex media can cause a pH-shift (peptone, yeast extract)
Basic material	pH sensor HP8-1427-02_3 (at least stable for 7 days with CertiPUR-buffer) pH sensors are light-sensitive; please protect them from direct light!

pH calibration

Buffer	CertiPUR Reference Material Buffer solutions Set (pH 3.00 ± 0.01 / pH 4.00 ± 0.015 / pH 9.00 ± 0.01 / pH 10.00 ± 0.03, 20 °C); 150 mM Na-Phosphate buffer (16 solutions)
Settings	BioLector protocol = pH-DO-calibration, T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = FlowerPlate (MTP-48-BOH)
Calibration device	Hardware ID: BL-02-000F-0033
Calibration phase offset	pH -1.83(pH Ser.3144-RD, gain 7)
Date of calibration	2017/04/04

EUROPE

m2p-labs GmbH
Arnold-Sommerfeld-Ring 2 | 52499 Baesweiler | Germany
Phone +49-2401-805-330 | Fax: +49-2401-805-333
info@m2p-labs.com | support@m2p-labs.com

USA

m2p-labs, Inc.
400 Oser Ave, Suite 1650 | Hauppauge, NY 11788 | USA
Phone +1-631-501-1878 | Fax +1-631-501-1060
infoUS@m2p-labs.com | supportUS@m2p-labs.com

Please enter these **calibration parameters** and the **Lot No.** into the BioLecture software!

DO calibration parameters Lot No. 1704 (BioLector® II/Pro, filter module ID-203)

Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
ϕ cal0	72.37	72.31	72.24	72.18	72.11	72.05	71.98
ϕ cal100	46.00	45.68	45.35	45.03	44.70	44.38	44.05
Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
ϕ cal0	71.92	71.85	71.78	71.72	71.65	71.59	71.52
ϕ cal100	43.73	43.40	43.08	42.75	42.43	42.10	41.78
Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
ϕ cal0	71.46	71.39	71.33	71.26	71.20	71.13	71.07
ϕ cal100	41.45	41.13	40.80	40.48	40.15	39.83	39.50

DO sensor properties

Dynamic range	0 - 100 % air saturation (a.s.)
Resolution	Up to 0.1 % O ₂ (software)
Accuracy	± 5% dissolved oxygen (batch calibration)
Drift at 0% oxygen	< 0.5% O ₂ per day (sampling interval of 6 min)
Response time (t90)	< 30 s
Temperature range	5 – 50°C
Sensor stability	sensor material can be degraded by some microorganisms
Cross-sensitivity to	Organic solvents, such as acetone, toluene, chloroform or methylene chloride, Chlorine gas; high concentration of fluorescent molecules in the visible range can interfere (mCherry, tdTomato, dsRed, Nile red); complex media can cause a DO-shift
Basic material	Oxygen sensor Pst3-HG-1426-03_3 (at least stable for 7 days with CertiPUR-buffer) DO sensors are light-sensitive; please protect them from direct light!

DO calibration

Calibration	0.5 M Sulfite system (Two-point calibration with oxygen-free environment (sodium sulfite) and air-saturated environment)
Settings	BioLector protocol = pH-DO-calibration, T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = FlowerPlate (MTP-48-BOH)
Calibration device	Hardware ID: BL-02-000F-0033
Calibration phase offset	DO -360.85 (DO Ser.4154-RD, gain 7)
Date of calibration	2017/04/04

Sterilization procedure

Sterilization	Gamma irradiation (15 kGy)
BGS-certificate No	353634
Date of sterilization	2017/03/26

EUROPE

m2p-labs GmbH
 Arnold-Sommerfeld-Ring 2 | 52499 Baesweiler | Germany
 Phone +49-2401-805-330 | Fax: +49-2401-805-333
 info@m2p-labs.com | support@m2p-labs.com

USA

m2p-labs, Inc.
 400 Oser Ave, Suite 1650 | Hauppauge, NY 11788 | USA
 Phone +1-631-501-1878 | Fax +1-631-501-1060
 infoUS@m2p-labs.com | supportUS@m2p-labs.com