

Please enter these **calibration parameters** and the **Lot No.** into the BioLecture software!

**pH calibration parameters Lot No. 1724 (BioLector® II/Pro, filter module ID-202)**

|                 |       |       |       |       |       |       |       |
|-----------------|-------|-------|-------|-------|-------|-------|-------|
| Temperature     | 20°C  | 21°C  | 22°C  | 23°C  | 24°C  | 25°C  | 26°C  |
| $\phi$ min      | 65.07 | 64.99 | 64.91 | 64.83 | 64.74 | 64.66 | 64.58 |
| $\phi$ max      | 19.06 | 19.06 | 19.06 | 19.05 | 19.05 | 19.05 | 19.05 |
| dpH             | 0.49  | 0.49  | 0.49  | 0.49  | 0.49  | 0.49  | 0.49  |
| pH <sub>0</sub> | 6.78  | 6.77  | 6.75  | 6.74  | 6.73  | 6.72  | 6.71  |
| Temperature     | 27°C  | 28°C  | 29°C  | 30°C  | 31°C  | 32°C  | 33°C  |
| $\phi$ min      | 64.49 | 64.41 | 64.33 | 64.25 | 64.16 | 64.08 | 64.00 |
| $\phi$ max      | 19.05 | 19.04 | 19.04 | 19.04 | 19.04 | 19.04 | 19.03 |
| dpH             | 0.49  | 0.49  | 0.49  | 0.49  | 0.49  | 0.49  | 0.49  |
| pH <sub>0</sub> | 6.70  | 6.69  | 6.68  | 6.67  | 6.66  | 6.65  | 6.63  |
| Temperature     | 34°C  | 35°C  | 36°C  | 37°C  | 38°C  | 39°C  | 40°C  |
| $\phi$ min      | 63.92 | 63.83 | 63.75 | 63.67 | 63.58 | 63.50 | 63.42 |
| $\phi$ max      | 19.03 | 19.03 | 19.03 | 19.03 | 19.02 | 19.02 | 19.02 |
| dpH             | 0.49  | 0.49  | 0.49  | 0.49  | 0.49  | 0.49  | 0.49  |
| pH <sub>0</sub> | 6.62  | 6.61  | 6.60  | 6.59  | 6.58  | 6.57  | 6.56  |

**pH sensor properties**

|                     |   |
|---------------------|---|
| Dynamic range       | pH 4.05 - 8.70  |
| Resolution          | Up to 0.01 pH (software)  |
| Accuracy            | ± 0.25 pH at pH 4.55 - 5.20; ± 0.1 pH at pH 5.20 – 7.60; ± 0.25 pH at pH 7.60 - 8.25 (batch calibration)  |
| Response time (t90) | At 25 °C < 30 s   |
| Drift at pH = 7     | < 0.005 pH per day (sampling interval of 6 min)   |
| Temperature range   | 5 °C to 50 °C   |
| Compatibility       | Aqueous solutions, ethanol, methanol (max. 5 % v/v)   |
| Sensor stability    | sensor material can be degraded by some microorganisms  |
| Cross-sensitivity   | Reduced to ionic strength (salinity);<br>high concentration of fluorescent molecules in the visible range can interfere (GFP, (e)YFP);<br>complex media can cause a pH-shift (peptone, yeast extract) |
| Basic material      | pH sensor HP8-1427-02_3 (at least stable for 7 days with CertiPUR-buffer)<br><b>pH sensors are light-sensitive; please protect them from direct light!</b>  |

**pH calibration**

|                          |  |
|--------------------------|--|
| Buffer                   | CertiPUR Reference Material Buffer solutions Set<br>(pH 3.00 ± 0.01 / pH 4.00 ± 0.015 / pH 9.00 ± 0.01 / pH 10.00 ± 0.03, 20 °C);<br>150 mM Na-Phosphate buffer (16 solutions) |
| Settings                 | BioLector protocol = pH-DO-calibration, T = 20-40 °C, 800 rpm, 1000 µL/well,<br>shaking diameter 3 mm, MTP-type = FlowerPlate (MTP-48-BOH)                                     |
| Calibration device       | Hardware ID: BL-02-000F-0032   |
| Calibration phase offset | pH -1.180 (pH Ser.3111-RD, gain 7)   |
| Date of calibration      | 2017/12/11   |

**EUROPE**

m2p-labs GmbH  
 Arnold-Sommerfeld-Ring 2 | 52499 Baesweiler | Germany  
 Phone +49-2401-805-330 | Fax: +49-2401-805-333  
 info@m2p-labs.com | support@m2p-labs.com

**USA**

m2p-labs, Inc.  
 400 Oser Ave, Suite 1650 | Hauppauge, NY 11788 | USA  
 Phone +1-631-501-1878 | Fax +1-631-501-1060  
 infoUS@m2p-labs.com | supportUS@m2p-labs.com

Please enter these **calibration parameters** and the **Lot No.** into the BioLecture software!

### DO calibration parameters Lot No. 1724 (BioLector® II/Pro, filter module ID-203)

|             |       |       |       |       |       |       |       |
|-------------|-------|-------|-------|-------|-------|-------|-------|
| Temperature | 20°C  | 21°C  | 22°C  | 23°C  | 24°C  | 25°C  | 26°C  |
| φ cal0      | 72.46 | 72.40 | 72.34 | 72.28 | 72.22 | 72.16 | 72.10 |
| φ cal100    | 45.41 | 45.16 | 44.91 | 44.66 | 44.41 | 44.16 | 43.91 |
| Temperature | 27°C  | 28°C  | 29°C  | 30°C  | 31°C  | 32°C  | 33°C  |
| φ cal0      | 72.04 | 71.99 | 71.93 | 71.87 | 71.81 | 71.75 | 71.69 |
| φ cal100    | 43.66 | 43.42 | 43.17 | 42.92 | 42.67 | 42.42 | 42.17 |
| Temperature | 34°C  | 35°C  | 36°C  | 37°C  | 38°C  | 39°C  | 40°C  |
| φ cal0      | 71.63 | 71.57 | 71.51 | 71.45 | 71.39 | 71.33 | 71.27 |
| φ cal100    | 41.92 | 41.68 | 41.43 | 41.18 | 40.93 | 40.68 | 40.43 |

### DO sensor properties

|                      |   |
|----------------------|---|
| Dynamic range        | 0 - 100 % air saturation (a.s.)   |
| Resolution           | Up to 0.1 % O <sub>2</sub> (software)   |
| Accuracy             | ± 5% dissolved oxygen (batch calibration)   |
| Drift at 0% oxygen   | < 0.5% O <sub>2</sub> per day (sampling interval of 6 min)  |
| Response time (t90)  | < 30 s  |
| Temperature range    | 5 – 50°C  |
| Sensor stability     | sensor material can be degraded by some microorganisms  |
| Cross-sensitivity to | Organic solvents, such as acetone, toluene, chloroform or methylene chloride, Chlorine gas; high concentration of fluorescent molecules in the visible range can interfere (mCherry, tdTomato, dsRed, Nile red); complex media can cause a DO-shift |
| Basic material       | Oxygen sensor Pst3-HG-1426-03_3 (at least stable for 7 days with CertiPUR-buffer)<br><b>DO sensors are light-sensitive; please protect them from direct light!</b>  |

### DO calibration

|                          |   |
|--------------------------|---|
| Calibration              | 0.5 M Sulfite system (Two-point calibration with oxygen-free environment (sodium sulfite) and air-saturated environment)                |
| Settings                 | BioLector protocol = pH-DO-calibration, T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = FlowerPlate (MTP-48-BOH) |
| Calibration device       | Hardware ID: BL-02-000F-0032  |
| Calibration phase offset | DO -360.59 (DO Ser.4103-RD, gain 7)   |
| Date of calibration      | 2017/12/11  |

### Sterilization procedure

|                       |                            |
|-----------------------|----------------------------|
| Sterilization         | Gamma irradiation (15 kGy) |
| BGS-certificate No    | 442942                     |
| Date of sterilization | 2017/12/04                 |

#### EUROPE

m2p-labs GmbH  
 Arnold-Sommerfeld-Ring 2 | 52499 Baesweiler | Germany  
 Phone +49-2401-805-330 | Fax: +49-2401-805-333  
 info@m2p-labs.com | support@m2p-labs.com

#### USA

m2p-labs, Inc.  
 400 Oser Ave, Suite 1650 | Hauppauge, NY 11788 | USA  
 Phone +1-631-501-1878 | Fax +1-631-501-1060  
 infoUS@m2p-labs.com | supportUS@m2p-labs.com