

Please enter these **calibration parameters** and the **Lot No.** into the BioLection software!

pH calibration parameters Lot No. 1805-S (BioLector® II/Pro, filter module ID-221)

Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
ϕ min	67.97	67.87	67.77	67.67	67.58	67.48	67.38
ϕ max	13.34	13.25	13.16	13.07	12.98	12.89	12.80
dpH	0.75	0.75	0.75	0.74	0.74	0.74	0.74
pH ₀	6.37	6.36	6.35	6.34	6.33	6.32	6.31
Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
ϕ min	67.29	67.19	67.09	66.99	66.90	66.80	66.70
ϕ max	12.71	12.63	12.54	12.45	12.36	12.27	12.18
dpH	0.74	0.74	0.74	0.74	0.73	0.73	0.73
pH ₀	6.30	6.28	6.27	6.26	6.25	6.24	6.23
Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
ϕ min	66.61	66.51	66.41	66.31	66.22	66.12	66.02
ϕ max	12.09	12.00	11.91	11.82	11.73	11.65	11.56
dpH	0.73	0.73	0.73	0.73	0.73	0.73	0.72
pH ₀	6.22	6.21	6.20	6.19	6.18	6.17	6.16

pH sensor properties

Dynamic range	pH 2.50 - 8.80
Resolution	Up to 0.01 pH (software)
Accuracy	± 0.25 pH at pH 3.50 - 5.70; ± 0.1 pH at pH 5.70 – 6.25; ± 0.25 pH at pH 6.25 – 8.15 (batch calibration)
Response time (t90)	At 25 °C < 30 s
Drift at pH = 7	< 0.005 pH per day (sampling interval of 6 min)
Temperature range	5 °C to 50 °C
Compatibility	Aqueous solutions, ethanol, methanol (max. 5 % v/v)
Sensor stability	sensor material can be degraded by some microorganisms
Cross-sensitivity	Reduced to ionic strength (salinity); high concentration of fluorescent molecules in the visible range can interfere (GFP, (e)YFP); complex media can cause a pH-shift (peptone, yeast extract)
Basic material	pH sensor LG1- 1737-01(at least stable for 7 days with CertiPUR-buffer) pH sensors are light-sensitive; please protect them from direct light!

pH calibration

Buffer	CertiPUR Reference Material Buffer solutions Set (pH 1.00 ± 0.01 / pH 2.00 ± 0.015 / pH 9.00 ± 0.01 / pH 10.00 ± 0.03, 20 °C); 150 mM Citrate-Na-Phosphate buffer (16 solutions)
Settings	BioLector protocol = LG1-RF-calibration, T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Microfluidic FlowerPlate (MTP-MF32-BOH)
Calibration device	Hardware ID: BL-02-000F-0032
Calibration phase offset	pH -360.78 (pH Ser.3188-RD, gain 8)
Date of calibration	2018/02/15

Sterilization procedure

Sterilization	Beta irradiation (20 kGy)
BGS-certificate No	469217
Date of sterilization	2018/02/13

EUROPE

m2p-labs GmbH
Arnold-Sommerfeld-Ring 2 | 52499 Baesweiler | Germany
Phone +49-2401-805-330 | Fax: +49-2401-805-333
info@m2p-labs.com | support@m2p-labs.com

USA

m2p-labs, Inc.
400 Oser Ave, Suite 1650 | Hauppauge, NY 11788 | USA
Phone +1-631-501-1878 | Fax +1-631-501-1060
infoUS@m2p-labs.com | supportUS@m2p-labs.com