

Please enter these **calibration parameters** and the **Lot No.** into the BioLecture software!

**pH calibration parameters Lot No. 1905 (BioLector® II/Pro. filter module ID-221/421)**

Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
φ min	64.83	64.82	64.81	64.79	64.78	64.77	64.75
φ max	6.71	6.70	6.68	6.67	6.65	6.63	6.62
dpH	0.74	0.74	0.74	0.74	0.74	0.74	0.74
pH <sub>0</sub>	5.80	5.80	5.79	5.79	5.79	5.78	5.78
Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
φ min	64.74	64.73	64.71	64.70	64.69	64.67	64.66
φ max	6.60	6.59	6.57	6.56	6.54	6.52	6.51
dpH	0.74	0.74	0.74	0.74	0.74	0.74	0.74
pH <sub>0</sub>	5.77	5.77	5.77	5.76	5.76	5.76	5.75
Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
φ min	64.65	64.63	64.62	64.61	64.59	64.58	64.56
φ max	6.49	6.48	6.46	6.45	6.43	6.41	6.40
dpH	0.74	0.74	0.74	0.74	0.74	0.74	0.74
pH <sub>0</sub>	5.75	5.75	5.74	5.74	5.74	5.73	5.73

**pH sensor properties**

Dynamic range	pH 3.10 - 7.95
Resolution	Up to 0.01 pH (software)
Accuracy	± 0.25 pH at pH 3.80 - 4.25; ± 0.1 pH at pH 4.25 – 6.80; ± 0.25 pH at pH 6.80 - 7.25 (batch calibration)
Response time (t90)	At 25 °C < 30 s
Drift at pH = 7	< 0.005 pH per day (sampling interval of 6 min)
Temperature range	5 °C to 50 °C
Compatibility	Aqueous solutions. ethanol. methanol (max. 5 % v/v)
Sensor stability	sensor material can be degraded by some microorganisms
Cross-sensitivity	Reduced to ionic strength (salinity); high concentration of fluorescent molecules in the visible range can interfere (GFP. (e)YFP); complex media can cause a pH-shift (peptone. yeast extract)
Basic material	pH sensor LG1-v1-1816-01 (at least stable for 7 days with CertiPUR-buffer) <b>pH sensors are light-sensitive; please protect them from direct light!</b>

**pH calibration**

Buffer	CertiPUR Reference Material Buffer solutions Set (pH 1.00 ± 0.01 / pH 2.00 ± 0.015 / pH 9.00 ± 0.01 / pH 10.00 ± 0.03. 20 °C); 150 mM Citrat-Na-Phosphate buffer (16 solutions)
Settings	BioLector protocol = LG1-RF-calibration. T = 20–40 °C. 800 rpm. 1000 µL/well. shaking diameter 3 mm. MTP-type = Microfluidic FlowerPlate (MTP-RMF32-BOH2)
Calibration device	Hardware ID: BL-02-000F-0032
Calibration phase offset	pH -360.31 (pH Ser.3188-RD. gain 8)
Date of calibration	2019/02/18

**HEADQUARTERS EUROPE**

m2p-labs GmbH  
Arnold-Sommerfeld-Ring 2  
52499 Baesweiler. Germany  
Tel.:+ 49 - 2401 805 330  
Fax:+49 - 2401 805 33  
info@m2p-labs.com

**USA / CANADA**

m2p-labs. Inc.  
400 Oser Ave. Suite 1650  
Hauppauge. NY 11788. USA  
Phone:+1 631 501 1878  
Fax:+1 631 501 1060  
infoUS@m2p-labs.com

**ASIA PACIFIC**

m2p-labs Limited  
Unit 117. Biotech Centre 2. HKSTP  
Shatin. NT. Hong Kong  
Phone:+852 6092 6778  
Fax:+852 3594 6381  
infoAsia@m2p-labs.com

Please enter these **calibration parameters** and the **Lot No.** into the BioLecture software!

### DO calibration parameters Lot No. 1905 (BioLector® II/Pro. filter module ID-228/428)

Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
ϕ cal0	68.22	68.21	68.19	68.18	68.17	68.16	68.15
ϕ cal100	42.08	41.98	41.88	41.78	41.67	41.57	41.47
Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
ϕ cal0	68.14	68.12	68.11	68.10	68.09	68.08	68.06
ϕ cal100	41.37	41.27	41.16	41.06	40.96	40.86	40.75
Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
ϕ cal0	68.05	68.04	68.03	68.02	68.00	67.99	67.98
ϕ cal100	40.65	40.55	40.45	40.35	40.24	40.14	40.04

### DO sensor properties

Dynamic range	0 - 100 % air saturation (a.s.)
Resolution	Up to 0.1 % O <sub>2</sub> (software)
Accuracy	± 5% dissolved oxygen (batch calibration)
Drift at 0% oxygen	< 0.5% O <sub>2</sub> per day (sampling interval of 6 min)
Response time (t90)	< 30 s
Temperature range	5 – 50°C
Sensor stability	sensor material can be degraded by some microorganisms
Cross-sensitivity to	Organic solvents, such as acetone, toluene, chloroform or methylene chloride. Chlorine gas; high concentration of fluorescent molecules in the visible range can interfere (mCherry, tdTomato, dsRed, Nile red); complex media can cause a DO-shift
Basic material	Oxygen sensor RF- 07/2018(at least stable for 7 days with CertiPUR-buffer) <b>DO sensors are light-sensitive; please protect them from direct light!</b>

### DO calibration

Calibration	0.5 M Sulfite system (Two-point calibration with oxygen-free environment (sodium sulfite) and air-saturated environment)
Settings	BioLector protocol = LG1-RF-calibration. T = 20-40 °C. 800 rpm. 1000 µL/well. shaking diameter 3 mm. MTP-type = Microfluidic FlowerPlate (MTP-RMF32-BOH2)
Calibration device	Hardware ID: BL-02-000F-0032
Calibration phase offset	DO -360.39 (DO Ser.4170-RD, gain 4)
Date of calibration	2019/02/18

### Sterilization procedure

Sterilization	Beta irradiation (20 kGy)
BGS-certificate No	58719
Date of sterilization	2019/01/31

#### HEADQUARTERS EUROPE

m2p-labs GmbH  
Arnold-Sommerfeld-Ring 2  
52499 Baesweiler, Germany  
Tel.: +49 - 2401 805 330  
Fax: +49 - 2401 805 33  
info@m2p-labs.com

#### USA / CANADA

m2p-labs, Inc.  
400 Oser Ave. Suite 1650  
Hauppauge, NY 11788, USA  
Phone: +1 631 501 1878  
Fax: +1 631 501 1060  
infoUS@m2p-labs.com

#### ASIA PACIFIC

m2p-labs Limited  
Unit 117, Biotech Centre 2, HKSTP  
Shatin, NT, Hong Kong  
Phone: +852 6092 6778  
Fax: +852 3594 6381  
infoAsia@m2p-labs.com