

Please enter these **calibration parameters** and the **Lot No.** into the BioLecture software!

**pH calibration parameters Lot No. 2010102 (BioLector® I, filter module ID-102/-302)**

| Temperature     | 20°C  | 21°C  | 22°C  | 23°C  | 24°C  | 25°C  | 26°C  |
|-----------------|-------|-------|-------|-------|-------|-------|-------|
| $\phi$ min      | 56.83 | 56.75 | 56.67 | 56.60 | 56.52 | 56.44 | 56.36 |
| $\phi$ max      | 11.52 | 11.52 | 11.51 | 11.51 | 11.51 | 11.51 | 11.51 |
| dpH             | 0.53  | 0.53  | 0.53  | 0.53  | 0.53  | 0.53  | 0.53  |
| pH <sub>0</sub> | 6.22  | 6.22  | 6.21  | 6.21  | 6.20  | 6.20  | 6.19  |
| Temperature     | 27°C  | 28°C  | 29°C  | 30°C  | 31°C  | 32°C  | 33°C  |
| $\phi$ min      | 56.28 | 56.21 | 56.13 | 56.05 | 55.97 | 55.89 | 55.81 |
| $\phi$ max      | 11.50 | 11.50 | 11.50 | 11.50 | 11.49 | 11.49 | 11.49 |
| dpH             | 0.53  | 0.53  | 0.53  | 0.53  | 0.54  | 0.54  | 0.54  |
| pH <sub>0</sub> | 6.18  | 6.18  | 6.17  | 6.17  | 6.16  | 6.15  | 6.15  |
| Temperature     | 34°C  | 35°C  | 36°C  | 37°C  | 38°C  | 39°C  | 40°C  |
| $\phi$ min      | 55.74 | 55.66 | 55.58 | 55.50 | 55.42 | 55.35 | 55.27 |
| $\phi$ max      | 11.49 | 11.49 | 11.48 | 11.48 | 11.48 | 11.48 | 11.48 |
| dpH             | 0.54  | 0.54  | 0.54  | 0.54  | 0.54  | 0.54  | 0.54  |
| pH <sub>0</sub> | 6.14  | 6.14  | 6.13  | 6.13  | 6.12  | 6.11  | 6.11  |

**pH sensor properties**

|                     |   |
|---------------------|---|
| Dynamic range       | pH 4.30-7.75  |
| Resolution          | Up to 0.01 pH (software)  |
| Accuracy            | ± 0.25 pH at pH 4.80 - 5.10; ± 0.1 pH at pH 5.10 – 6.95; ± 0.25 pH at pH 6.95 - 7.25 (batch calibration)  |
| Response time (t90) | At 25 °C < 30 s   |
| Drift at pH = 7     | < 0.005 pH per day (sampling interval of 6 min)   |
| Temperature range   | 5 °C to 50 °C   |
| Compatibility       | Aqueous solutions, ethanol, methanol (max. 5 % v/v)   |
| Sensor stability    | sensor material can be degraded by some microorganisms  |
| Cross-sensitivity   | Reduced to ionic strength (salinity);<br>high concentration of fluorescent molecules in the visible range can interfere (GFP, (e)YFP);<br>complex media can cause a pH-shift (peptone, yeast extract) |
| Basic material      | pH sensor HP8-1811-01_3 (at least stable for 7 days with CertiPUR-buffer)<br><b>pH sensors are light-sensitive; please protect them from direct light!</b>  |

**pH calibration**

|                          |   |
|--------------------------|---|
| Buffer                   | CertiPUR Reference Material Buffer solutions Set<br>(pH 3.00 ± 0.01 / pH 4.00 ± 0.015 / pH 9.00 ± 0.01 / pH 10.00 ± 0.03, 20 °C);<br>150 mM Citrat-Na-Phosphate buffer (16 solutions) |
| Settings                 | BioLector protocol = pH-DO-calibration, T = 20-40 °C, 800 rpm, 1000 µL/well,<br>shaking diameter 3 mm, MTP-type = Round Plate (MTP-R48-BOH1)  |
| Calibration device       | Hardware ID: BL092-CX-4A7394  |
| Calibration phase offset | pH 255.9 (pH Ser. 3403, gain 55)  |
| Date of calibration      | 2020/10/22  |

**HEADQUARTERS EUROPE**

m2p-labs GmbH Phone +49 - 2401 805 330  
Arnold-Sommerfeld-Ring 2 Fax +49 - 2401 805 33  
52499 Baesweiler, Germany info@m2p-labs.com

**SUPPORT**

**EUROPE**  
Phone +49 - 2401 805 335  
support@m2p-labs.com

**AMERICA**  
Phone +1 631 501 1878  
supportUS@m2p-labs.com

**ASIA PACIFIC**  
Phone +852 6092 6778  
supportAsia@m2p-labs.com

Please enter these **calibration parameters** and the **Lot No.** into the BioLecture software!

### DO calibration parameters Lot No. 2010102 (BioLector® I, filter module ID-103/-303)

|             |       |       |       |       |       |       |       |
|-------------|-------|-------|-------|-------|-------|-------|-------|
| Temperature | 20°C  | 21°C  | 22°C  | 23°C  | 24°C  | 25°C  | 26°C  |
| φ cal0      | 72.30 | 72.21 | 72.12 | 72.03 | 71.95 | 71.86 | 71.77 |
| φ cal100    | 41.67 | 41.42 | 41.17 | 40.92 | 40.67 | 40.42 | 40.18 |
| Temperature | 27°C  | 28°C  | 29°C  | 30°C  | 31°C  | 32°C  | 33°C  |
| φ cal0      | 71.68 | 71.59 | 71.50 | 71.41 | 71.32 | 71.23 | 71.14 |
| φ cal100    | 39.93 | 39.68 | 39.43 | 39.18 | 38.93 | 38.68 | 38.44 |
| Temperature | 34°C  | 35°C  | 36°C  | 37°C  | 38°C  | 39°C  | 40°C  |
| φ cal0      | 71.05 | 70.96 | 70.87 | 70.78 | 70.69 | 70.60 | 70.51 |
| φ cal100    | 38.19 | 37.94 | 37.69 | 37.44 | 37.19 | 36.94 | 36.70 |

### DO sensor properties

|                                  |   |
|----------------------------------|---|
| Dynamic range                    | 0 - 100 % air saturation (a.s.)   |
| Resolution                       | Up to 0.5 % O <sub>2</sub> (software)   |
| Accuracy                         | ± 5% dissolved oxygen (batch calibration)   |
| Drift at 0% oxygen               | < 0.5% O <sub>2</sub> per day (sampling interval of 6 min)  |
| Response time (t <sub>90</sub> ) | < 30 s  |
| Temperature range                | 5 – 50°C  |
| Sensor stability                 | sensor material can be degraded by some microorganisms  |
| Cross-sensitivity to             | Organic solvents, such as acetone, toluene, chloroform or methylene chloride, Chlorine gas; high concentration of fluorescent molecules in the visible range can interfere (mCherry, tdTomato, dsRed, Nile red); complex media can cause a DO-shift |
| Basic material                   | Oxygen sensor Pst3-HG-1810-01_3 (at least stable for 7 days with CertiPUR-buffer)<br><b>DO sensors are light-sensitive; please protect them from direct light!</b>  |

### DO calibration

|                          |   |
|--------------------------|---|
| Calibration              | 0.5 M Sulfite system (Two-point calibration with oxygen-free environment (sodium sulfite) and air-saturated environment)                  |
| Settings                 | BioLector protocol = pH-DO-calibration, T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Round Plate (MTP-R48-BOH1) |
| Calibration device       | Hardware ID: BL092-CX-4A7394  |
| Calibration phase offset | DO 332.5 (DO Ser.3402, gain 70)   |
| Date of calibration      | 2020/10/22  |

### Sterilization procedure

|                       |                           |
|-----------------------|---------------------------|
| Sterilization         | Beta irradiation (20 kGy) |
| BGS-certificate No    | 815113                    |
| Date of sterilization | 2020/10/15                |

#### HEADQUARTERS EUROPE

m2p-labs GmbH  
Arnold-Sommerfeld-Ring 2  
52499 Baesweiler, Germany  
Phone +49 - 2401 805 330  
Fax +49 - 2401 805 33  
info@m2p-labs.com

#### SUPPORT

**EUROPE**  
Phone +49 - 2401 805 335  
support@m2p-labs.com

**AMERICA**  
Phone +1 631 501 1878  
supportUS@m2p-labs.com

**ASIA PACIFIC**  
Phone +852 6092 6778  
supportAsia@m2p-labs.com