

Please enter these **calibration parameters** and the **Lot No.** into the BioLector software!

**pH calibration parameters Lot No.2102121 (BioLector® Pro, filter module ID-202/-402)**

Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
φ min	63.07	63.02	62.97	62.92	62.87	62.82	62.77
φ max	13.20	13.21	13.22	13.23	13.24	13.25	13.26
dpH	0.57	0.57	0.57	0.57	0.57	0.57	0.57
pH <sub>0</sub>	6.24	6.24	6.23	6.22	6.21	6.21	6.20

Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
φ min	62.71	62.66	62.61	62.56	62.51	62.46	62.41
φ max	13.27	13.28	13.29	13.30	13.31	13.32	13.33
dpH	0.57	0.57	0.57	0.57	0.57	0.57	0.57
pH <sub>0</sub>	6.19	6.18	6.18	6.17	6.16	6.15	6.15

Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
φ min	62.36	62.31	62.26	62.21	62.16	62.11	62.06
φ max	13.35	13.36	13.37	13.38	13.39	13.40	13.41
dpH	0.57	0.57	0.57	0.57	0.57	0.57	0.58
pH <sub>0</sub>	6.14	6.13	6.12	6.12	6.11	6.10	6.09

**pH sensor properties**

Dynamic range	pH 4.20 - 7.90
Resolution	Up to 0.01 pH (software)
Accuracy	± 0.25 pH at pH 4.70-5.05; ± 0.1 pH at pH 5.05-7.00; ± 0.25 pH at pH 7.00-7.35 batch calibration
Response time (t90)	At 25 °C < 30 s
Drift at pH = 7	< 0.005 pH per day (sampling interval of 6 min)
Temperature range	5 °C to 50 °C
Compatibility	Aqueous solutions, ethanol, methanol (max. 5 % v/v)
Sensor stability	Sensor material can be degraded by some microorganisms
Cross-sensitivity	Reduced to ionic strength (salinity); high concentration of fluorescent molecules in the visible range can interfere (GFP, (e)YFP); complex media can cause a pH-shift (peptone, yeast extract)
Basic material	pH sensor HP8-1811-01_3 (at least stable for 7 days with CertiPUR-buffer) <b>pH sensors are light-sensitive; please protect them from direct light!</b>

**pH calibration**

Buffer	CertiPUR Reference Material Buffer solutions Set (pH 3.00 ± 0.01 / pH 4.00 ± 0.015 / pH 9.00 ± 0.01 / pH 10.00 ± 0.03, 20 °C); 150 mM Na-Phosphate buffer (16 solutions)
Settings	BioLector protocol = HP8-PSt3-calibration, T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Microfluidic Flower (MTP-MF32C-BOH1)
Calibration device	Hardware ID: BL-09-000F-0032
Calibration phase offset	pH -1.46 (pH Ser. 3111, gain 7)
Date of calibration	2021-03-15

**HEADQUARTERS EUROPE**

m2p-labs GmbH  
Arnold-Sommerfeld-Ring 2  
52499 Baesweiler, Germany  
Phone +49 -2401 805 330  
Fax +49 -2401 805 33  
info@m2p-labs.com

**SUPPORT**

**EUROPE**  
Phone +49 - 2401 805 335  
support@m2p-labs.com

**N./S. AMERICAS**  
Phone +1 631 501 1878  
supportUS@m2p-labs.com

**ASIA PACIFIC**  
Phone +852 6092 6778  
supportAsia@m2p-labs.com

Please enter these **calibration parameters** and the **Lot No.** into the BioLecton software!

**DO calibration parameters Lot No.2102121 (BioLector® Pro, filter module ID-203/-403)**

Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
ϕ cal0	72.93	72.88	72.84	72.80	72.76	72.71	72.67
ϕ cal100	43.43	43.20	42.98	42.75	42.53	42.30	42.08

Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
ϕ cal0	72.63	72.59	72.54	72.50	72.46	72.42	72.37
ϕ cal100	41.86	41.63	41.41	41.18	40.96	40.73	40.51

Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
ϕ cal0	72.33	72.29	72.25	72.20	72.16	72.12	72.08
ϕ cal100	40.28	40.06	39.83	39.61	39.39	39.16	38.94

**DO sensor properties**

Dynamic range	0 - 100 % air saturation (a.s.)
Resolution	Up to 0.1 % O <sub>2</sub> (software)
Accuracy	± 5% dissolved oxygen (batch calibration)
Drift at 0% oxygen	< 0.5% O <sub>2</sub> per day (sampling interval of 6 min)
Response time (t <sub>90</sub> )	< 30 s
Temperature range	5 – 50°C
Sensor stability	Sensor material can be degraded by some microorganisms
Cross-sensitivity to	Organic solvents, such as acetone, toluene, chloroform or methylene chloride, Chlorine gas; high concentration of fluorescent molecules in the visible range can interfere (mCherry, tdTomato, dsRed, Nile red); complex media can cause a DO-shift
Basic material	Oxygen sensor PSt3-HG-1810-01_3 (at least stable for 7 days with CertiPUR-buffer) <b>DO sensors are light-sensitive; please protect them from direct light!</b>

**DO calibration**

Calibration	0.5 M Sulfite system (Two-point calibration with oxygen-free environment (sodium sulfite) and air-saturated environment)
Settings	BioLector protocol = HP8-PSt3-calibration, T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Microfluidic Flower (MTP-MF32C-BOH1)
Calibration device	Hardware ID: BL-09-000F-0032
Calibration phase offset	DO -360.26 (DO Ser. 4103, gain 7)
Date of calibration	2021-03-15

**Sterilization procedure**

Sterilization	Beta irradiation (20 kGy)
BGS-certificate No	867189
Date of sterilization	2021-03-03

**HEADQUARTERS EUROPE**

m2p-labs GmbH  
Arnold-Sommerfeld-Ring 2  
52499 Baesweiler, Germany  
Phone +49 -2401 805 330  
Fax +49 -2401 805 33  
info@m2p-labs.com

**SUPPORT**

**EUROPE**  
Phone +49 - 2401 805 335  
support@m2p-labs.com

**N./S. AMERICAS**  
Phone +1 631 501 1878  
supportUS@m2p-labs.com

**ASIA PACIFIC**  
Phone +852 6092 6778  
supportAsia@m2p-labs.com