

Please enter these **calibration parameters** and the **Lot No.** into the BioLector software!

pH calibration parameters Lot No.2106221 (BioLector® Pro, filter module ID-221/-421)

Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
φ min	66.76	66.71	66.65	66.59	66.54	66.48	66.42
φ max	13.41	13.34	13.27	13.20	13.13	13.06	12.99
dpH	0.82	0.82	0.82	0.82	0.82	0.82	0.82
pH ₀	6.63	6.62	6.61	6.60	6.60	6.59	6.58

Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
φ min	66.37	66.31	66.25	66.20	66.14	66.08	66.03
φ max	12.92	12.85	12.78	12.71	12.64	12.57	12.50
dpH	0.82	0.82	0.82	0.82	0.82	0.82	0.82
pH ₀	6.57	6.56	6.56	6.55	6.54	6.53	6.52

Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
φ min	65.97	65.91	65.86	65.80	65.74	65.69	65.63
φ max	12.44	12.37	12.30	12.23	12.16	12.09	12.02
dpH	0.82	0.82	0.82	0.82	0.82	0.82	0.82
pH ₀	6.52	6.51	6.50	6.49	6.48	6.48	6.47

pH sensor properties

Dynamic range	pH 3.90 - 8.85
Resolution	Up to 0.01 pH (software)
Accuracy	± 0.25 pH at pH 4.70-5.25; ± 0.1 pH at pH 5.25-7.55; ± 0.25 pH at pH 7.55-8.05 batch calibration
Response time (t90)	At 25 °C < 30 s
Drift at pH = 7	< 0.005 pH per day (sampling interval of 6 min)
Temperature range	5 °C to 50 °C
Compatibility	Aqueous solutions, ethanol, methanol (max. 5 % v/v)
Sensor stability	Sensor material can be degraded by some microorganisms
Cross-sensitivity	Reduced to ionic strength (salinity); high concentration of fluorescent molecules in the visible range can interfere (GFP, (e)YFP); complex media can cause a pH-shift (peptone, yeast extract)
Basic material	pH sensor LG1 1939-01_2 (at least stable for 7 days with CertiPUR-buffer) pH sensors are light-sensitive; please protect them from direct light!

pH calibration

Buffer	CertiPUR Reference Material Buffer solutions Set (pH 2.00 ± 0.01 / pH 3.00 ± 0.015 / pH 9.00 ± 0.01 / pH 10.00 ± 0.03, 20 °C); 150 mM Citrat-Na-Phosphate buffer (16 solutions)
Settings	BioLector protocol = LG1-RF-calibration, T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Microfluidic Flower Plate (MTP-MF32C-BOH2)
Calibration device	Hardware ID: BL-09-000F-0032
Calibration phase offset	pH -360.15 (pH Ser. 3305, gain 8)
Date of calibration	2021-06-01

HEADQUARTERS EUROPE

m2p-labs GmbH
Arnold-Sommerfeld-Ring 2
52499 Baesweiler, Germany
Phone +49 -2401 805 330
Fax +49 -2401 805 33
info@m2p-labs.com

SUPPORT

EUROPE
Phone +49 - 2401 805 335
support@m2p-labs.com

N./S. AMERICAS
Phone +1 631 501 1878
supportUS@m2p-labs.com

ASIA PACIFIC
Phone +852 6092 6778
supportAsia@m2p-labs.com

Please enter these **calibration parameters** and the **Lot No.** into the BioLecton software!

DO calibration parameters Lot No.2106221 (BioLector® Pro, filter module ID-228/-428)

Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
ϕ cal0	72.06	72.05	72.03	72.02	72.01	71.99	71.98
ϕ cal100	42.12	41.95	41.79	41.63	41.47	41.31	41.15

Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
ϕ cal0	71.97	71.95	71.94	71.92	71.91	71.90	71.88
ϕ cal100	40.98	40.82	40.66	40.50	40.34	40.18	40.02

Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
ϕ cal0	71.87	71.85	71.84	71.83	71.81	71.80	71.78
ϕ cal100	39.85	39.69	39.53	39.37	39.21	39.05	38.88

DO sensor properties

Dynamic range	0 - 100 % air saturation (a.s.)
Resolution	Up to 0.1 % O ₂ (software)
Accuracy	± 5% dissolved oxygen (batch calibration)
Drift at 0% oxygen	< 0.5% O ₂ per day (sampling interval of 6 min)
Response time (t ₉₀)	< 30 s
Temperature range	5 – 50°C
Sensor stability	Sensor material can be degraded by some microorganisms
Cross-sensitivity to	Organic solvents, such as acetone, toluene, chloroform or methylene chloride, Chlorine gas; high concentration of fluorescent molecules in the visible range can interfere (mCherry, tdTomato, dsRed, Nile red); complex media can cause a DO-shift
Basic material	Oxygen sensor RF-210250002 (at least stable for 7 days with CertiPUR-buffer) DO sensors are light-sensitive; please protect them from direct light!

DO calibration

Calibration	0.5 M Sulfite system (Two-point calibration with oxygen-free environment (sodium sulfite) and air-saturated environment)
Settings	BioLector protocol = LG1-RF-calibration, T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Microfluidic Flower Plate (MTP-MF32C-B0H2)
Calibration device	Hardware ID: BL-09-000F-0032
Calibration phase offset	DO -360.44 (DO Ser. 4302-RD, gain 4)
Date of calibration	2021-06-01

Sterilization procedure

Sterilization	Beta irradiation (20 kGy)
BGS-certificate No	899513
Date of sterilization	2021-05-26

HEADQUARTERS EUROPE

m2p-labs GmbH
Arnold-Sommerfeld-Ring 2
52499 Baesweiler, Germany
Phone +49 -2401 805 330
Fax +49 -2401 805 33
info@m2p-labs.com

SUPPORT

EUROPE
Phone +49 - 2401 805 335
support@m2p-labs.com

N./S. AMERICAS
Phone +1 631 501 1878
supportUS@m2p-labs.com

ASIA PACIFIC
Phone +852 6092 6778
supportAsia@m2p-labs.com