

Please enter these **calibration parameters** and the **Lot No.** into the BioLector software!

**pH calibration parameters Lot No.2112222 (BioLector® Pro, filter module ID-221/-421)**

Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
φ min	67.01	66.96	66.91	66.87	66.82	66.77	66.72
φ max	12.94	12.89	12.83	12.78	12.73	12.68	12.63
dpH	0.79	0.79	0.79	0.78	0.78	0.78	0.78
pH <sub>0</sub>	6.62	6.61	6.61	6.60	6.59	6.59	6.58

Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
φ min	66.67	66.62	66.57	66.52	66.47	66.42	66.37
φ max	12.58	12.53	12.48	12.43	12.37	12.32	12.27
dpH	0.78	0.78	0.78	0.78	0.78	0.78	0.78
pH <sub>0</sub>	6.57	6.57	6.56	6.56	6.55	6.54	6.54

Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
φ min	66.33	66.28	66.23	66.18	66.13	66.08	66.03
φ max	12.22	12.17	12.12	12.07	12.02	11.97	11.91
dpH	0.78	0.78	0.78	0.78	0.78	0.78	0.78
pH <sub>0</sub>	6.53	6.52	6.52	6.51	6.50	6.50	6.49

**pH sensor properties**

Dynamic range	pH 3.95 - 8.75
Resolution	Up to 0.01 pH (software)
Accuracy	± 0.25 pH at pH 4.70-5.20; ± 0.1 pH at pH 5.20-7.55; ± 0.25 pH at pH 7.55-8.05 batch calibration
Response time (t90)	At 25 °C < 30 s
Drift at pH = 7	< 0.005 pH per day (sampling interval of 6 min)
Temperature range	5 °C to 50 °C
Compatibility	Aqueous solutions, ethanol, methanol (max. 5 % v/v)
Sensor stability	Sensor material can be degraded by some microorganisms
Cross-sensitivity	Reduced to ionic strength (salinity); high concentration of fluorescent molecules in the visible range can interfere (GFP, (e)YFP); complex media can cause a pH-shift (peptone, yeast extract)
Basic material	pH sensor LG1-1939-01_2 (at least stable for 7 days with CertiPUR-buffer) <b>pH sensors are light-sensitive; please protect them from direct light!</b>

**pH calibration**

Buffer	CertiPUR Reference Material Buffer solutions Set (pH 2.00 ± 0.01 / pH 3.00 ± 0.015 / pH 9.00 ± 0.01 / pH 10.00 ± 0.03, 20 °C); 150 mM Citrat-Na-Phosphate buffer (16 solutions)
Settings	BioLector protocol = LG1-RF-calibration, T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Microfluidic Round Well Plate (MTP-RMF32-BOH2)
Calibration device	Hardware ID: BL-09-000F-0032
Calibration phase offset	pH -360.15 (pH Ser. 3305, gain 8)
Date of calibration	2021-08-26

**HEADQUARTERS EUROPE**

m2p-labs GmbH  
Arnold-Sommerfeld-Ring 2  
52499 Baesweiler, Germany  
Phone +49 -2401 805 330  
Fax +49 -2401 805 33  
info@m2p-labs.com

**SUPPORT**

**EUROPE**  
Phone +49 - 2401 805 335  
support@m2p-labs.com

**N./S. AMERICAS**  
Phone +1 631 501 1878  
supportUS@m2p-labs.com

**ASIA PACIFIC**  
Phone +852 6092 6778  
supportAsia@m2p-labs.com

Please enter these **calibration parameters** and the **Lot No.** into the BioLector software!

**DO calibration parameters Lot No.2112222 (BioLector® Pro, filter module ID-228/-428)**

Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
ϕ cal0	71.39	71.38	71.38	71.37	71.37	71.36	71.36
ϕ cal100	40.80	40.67	40.54	40.42	40.29	40.17	40.04

Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
ϕ cal0	71.35	71.35	71.34	71.34	71.33	71.33	71.33
ϕ cal100	39.91	39.79	39.66	39.54	39.41	39.28	39.16

Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
ϕ cal0	71.32	71.32	71.31	71.31	71.30	71.30	71.29
ϕ cal100	39.03	38.91	38.78	38.65	38.53	38.40	38.28

**DO sensor properties**

Dynamic range	0 - 100 % air saturation (a.s.)
Resolution	Up to 0.1 % O <sub>2</sub> (software)
Accuracy	± 5% dissolved oxygen (batch calibration)
Drift at 0% oxygen	< 0.5% O <sub>2</sub> per day (sampling interval of 6 min)
Response time (t <sub>90</sub> )	< 30 s
Temperature range	5 – 50°C
Sensor stability	Sensor material can be degraded by some microorganisms
Cross-sensitivity to	Organic solvents, such as acetone, toluene, chloroform or methylene chloride, Chlorine gas; high concentration of fluorescent molecules in the visible range can interfere (mCherry, tdTomato, dsRed, Nile red); complex media can cause a DO-shift
Basic material	Oxygen sensor RF-21160295 (at least stable for 7 days with CertiPUR-buffer) <b>DO sensors are light-sensitive; please protect them from direct light!</b>

**DO calibration**

Calibration	0.5 M Sulfite system (Two-point calibration with oxygen-free environment (sodium sulfite) and air-saturated environment)
Settings	BioLector protocol = LG1-RF-calibration, T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Microfluidic Round Well Plate (MTP-RMF32-BOH2)
Calibration device	Hardware ID: BL-09-000F-0032
Calibration phase offset	DO -360.44 (DO Ser. 4302-RD, gain 4)
Date of calibration	2021-08-26

**Sterilization procedure**

Sterilization	Beta irradiation (20 kGy)
BGS-certificate No	932566
Date of sterilization	2021-08-17

**HEADQUARTERS EUROPE**

m2p-labs GmbH  
Arnold-Sommerfeld-Ring 2  
52499 Baesweiler, Germany  
Phone +49 -2401 805 330  
Fax +49 -2401 805 33  
info@m2p-labs.com

**SUPPORT**

**EUROPE**  
Phone +49 - 2401 805 335  
support@m2p-labs.com

**N./S. AMERICAS**  
Phone +1 631 501 1878  
supportUS@m2p-labs.com

**ASIA PACIFIC**  
Phone +852 6092 6778  
supportAsia@m2p-labs.com