

Please enter these **calibration parameters** and the **Lot No.** into the BioLecton software!

**pH calibration parameters Lot No.2114101+2114181 (BioLector® I, filter module ID-102/-302)**

Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
φ min	57.35	57.26	57.18	57.10	57.02	56.94	56.85
φ max	11.18	11.18	11.18	11.18	11.17	11.17	11.17
dpH	0.57	0.57	0.57	0.57	0.57	0.58	0.58
pH <sub>0</sub>	6.17	6.16	6.15	6.15	6.14	6.14	6.13

Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
φ min	56.77	56.69	56.61	56.52	56.44	56.36	56.28
φ max	11.17	11.17	11.16	11.16	11.16	11.16	11.15
dpH	0.58	0.58	0.58	0.58	0.58	0.58	0.58
pH <sub>0</sub>	6.12	6.12	6.11	6.10	6.10	6.09	6.09

Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
φ min	56.19	56.11	56.03	55.95	55.87	55.78	55.70
φ max	11.15	11.15	11.15	11.15	11.14	11.14	11.14
dpH	0.58	0.58	0.58	0.58	0.58	0.58	0.58
pH <sub>0</sub>	6.08	6.07	6.07	6.06	6.05	6.05	6.04

### pH sensor properties

Dynamic range	pH 4.10 - 7.80
Resolution	Up to 0.01 pH (software)
Accuracy	± 0.25 pH at pH 4.65-5.00; ± 0.1 pH at pH 5.00-6.90; ± 0.25 pH at pH 6.90-7.25 batch calibration
Response time (t90)	At 25 °C < 30 s
Drift at pH = 7	< 0.005 pH per day (sampling interval of 6 min)
Temperature range	5 °C to 50 °C
Compatibility	Aqueous solutions, ethanol, methanol (max. 5 % v/v)
Sensor stability	Sensor material can be degraded by some microorganisms
Cross-sensitivity	Reduced to ionic strength (salinity); high concentration of fluorescent molecules in the visible range can interfere (GFP, (e)YFP); complex media can cause a pH-shift (peptone, yeast extract)
Basic material	pH sensor HP8-1811-01_5 (at least stable for 7 days with CertiPUR-buffer) <b>pH sensors are light-sensitive; please protect them from direct light!</b>

### pH calibration

Buffer	CertiPUR Reference Material Buffer solutions Set (pH 3.00 ± 0.02 / pH 4.00 ± 0.02 / pH 9.00 ± 0.03 / pH 10.00 ± 0.03, 20 °C); 150 mM Na-Phosphate buffer (16 solutions)
Settings	BioLector protocol = pH-DO-calibration, T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Flower Plate (MTP-48-BOH1)
Calibration device	Hardware ID: BL092-CX-4A7394
Calibration phase offset	pH 255.90 (pH Ser. 3403, gain 55)
Date of calibration	2021-10-12

#### HEADQUARTERS EUROPE

m2p-labs GmbH  
Arnold-Sommerfeld-Ring 2  
52499 Baesweiler, Germany  
Phone +49 -2401 805 330  
Fax +49 -2401 805 33  
info@m2p-labs.com

#### SUPPORT

**EUROPE**  
Phone +49 - 2401 805 335  
support@m2p-labs.com

**N./S. AMERICAS**  
Phone +1 631 501 1878  
supportUS@m2p-labs.com

**ASIA PACIFIC**  
Phone +852 6092 6778  
supportAsia@m2p-labs.com

Please enter these **calibration parameters** and the **Lot No.** into the BioLector software!

**DO calibration parameters Lot No.2114101+2114181 (BioLector® I, filter module ID-103/-303)**

Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
ϕ cal0	73.64	73.56	73.47	73.39	73.31	73.22	73.14
ϕ cal100	41.82	41.66	41.50	41.33	41.17	41.00	40.84

Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
ϕ cal0	73.05	72.97	72.89	72.80	72.72	72.63	72.55
ϕ cal100	40.68	40.51	40.35	40.19	40.02	39.86	39.69

Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
ϕ cal0	72.47	72.38	72.30	72.21	72.13	72.04	71.96
ϕ cal100	39.53	39.37	39.20	39.04	38.88	38.71	38.55

### DO sensor properties

Dynamic range	0 - 100 % air saturation (a.s.)
Resolution	Up to 0.5 % O <sub>2</sub> (software)
Accuracy	± 5% dissolved oxygen (batch calibration)
Drift at 0% oxygen	< 0.5% O <sub>2</sub> per day (sampling interval of 6 min)
Response time (t <sub>90</sub> )	< 30 s
Temperature range	5 – 50°C
Sensor stability	Sensor material can be degraded by some microorganisms
Cross-sensitivity to	Organic solvents, such as acetone, toluene, chloroform or methylene chloride, Chlorine gas; high concentration of fluorescent molecules in the visible range can interfere (mCherry, tdTomato, dsRed, Nile red); complex media can cause a DO-shift
Basic material	Oxygen sensor PSt3-HG-1921-01_2 (at least stable for 7 days with CertiPUR-buffer) <b>DO sensors are light-sensitive; please protect them from direct light!</b>

### DO calibration

Calibration	0.5 M Sulfite system (Two-point calibration with oxygen-free environment (sodium sulfite) and air-saturated environment)
Settings	BioLector protocol = pH-DO-calibration, T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Flower Plate (MTP-48-BOH1)
Calibration device	Hardware ID: BL092-CX-4A7394
Calibration phase offset	DO 332.50 (DO Ser. 3402, gain 70)
Date of calibration	2021-10-12

### Sterilization procedure

Sterilization	Beta irradiation (20 kGy)
BGS-certificate No	952124
Date of sterilization	2021-10-05

#### HEADQUARTERS EUROPE

m2p-labs GmbH  
Arnold-Sommerfeld-Ring 2  
52499 Baesweiler, Germany  
Phone +49 -2401 805 330  
Fax +49 -2401 805 33  
info@m2p-labs.com

#### SUPPORT

**EUROPE**  
Phone +49 - 2401 805 335  
support@m2p-labs.com

**N./S. AMERICAS**  
Phone +1 631 501 1878  
supportUS@m2p-labs.com

**ASIA PACIFIC**  
Phone +852 6092 6778  
supportAsia@m2p-labs.com