

Please enter these **calibration parameters** and the **Lot No.** into the BioLector software!

pH calibration parameters Lot No.2114211 (BioLector® Pro, filter module ID-221/-421)

Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
φ min	69.65	69.55	69.44	69.34	69.24	69.13	69.03
φ max	14.18	14.09	14.00	13.92	13.83	13.74	13.65
dpH	0.74	0.74	0.74	0.74	0.74	0.74	0.74
pH ₀	6.58	6.58	6.57	6.56	6.55	6.55	6.54

Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
φ min	68.92	68.82	68.71	68.61	68.50	68.40	68.29
φ max	13.56	13.48	13.39	13.30	13.21	13.13	13.04
dpH	0.74	0.74	0.74	0.74	0.74	0.74	0.74
pH ₀	6.53	6.53	6.52	6.51	6.50	6.50	6.49

Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
φ min	68.19	68.08	67.98	67.88	67.77	67.67	67.56
φ max	12.95	12.86	12.78	12.69	12.60	12.51	12.43
dpH	0.74	0.74	0.74	0.74	0.74	0.74	0.74
pH ₀	6.48	6.48	6.47	6.46	6.45	6.45	6.44

pH sensor properties

Dynamic range	pH 4.00 - 8.60
Resolution	Up to 0.01 pH (software)
Accuracy	± 0.25 pH at pH 4.70-5.15; ± 0.1 pH at pH 5.15-7.50; ± 0.25 pH at pH 7.50-7.95 batch calibration
Response time (t90)	At 25 °C < 30 s
Drift at pH = 7	< 0.005 pH per day (sampling interval of 6 min)
Temperature range	5 °C to 50 °C
Compatibility	Aqueous solutions, ethanol, methanol (max. 5 % v/v)
Sensor stability	Sensor material can be degraded by some microorganisms
Cross-sensitivity	Reduced to ionic strength (salinity); high concentration of fluorescent molecules in the visible range can interfere (GFP, (e)YFP); complex media can cause a pH-shift (peptone, yeast extract)
Basic material	pH sensor LG1-1939-01_2 (at least stable for 7 days with CertiPUR-buffer) pH sensors are light-sensitive; please protect them from direct light!

pH calibration

Buffer	CertiPUR Reference Material Buffer solutions Set (pH 2.00 ± 0.01 / pH 3.00 ± 0.015 / pH 9.00 ± 0.01 / pH 10.00 ± 0.03, 20 °C); 150 mM Citrat-Na-Phosphate buffer (16 solutions)
Settings	BioLector protocol = LG1-RF-calibration, T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Microfluidic Flower Plate (MTP-MF32-BOH2)
Calibration device	Hardware ID: BL-09-000F-0032
Calibration phase offset	pH -360.15 (pH Ser. 3305, gain 8)
Date of calibration	2021-10-22

HEADQUARTERS EUROPE

m2p-labs GmbH
Arnold-Sommerfeld-Ring 2
52499 Baesweiler, Germany
Phone +49 -2401 805 330
Fax +49 -2401 805 33
info@m2p-labs.com

SUPPORT

EUROPE
Phone +49 - 2401 805 335
support@m2p-labs.com

N./S. AMERICAS
Phone +1 631 501 1878
supportUS@m2p-labs.com

ASIA PACIFIC
Phone +852 6092 6778
supportAsia@m2p-labs.com

Please enter these **calibration parameters** and the **Lot No.** into the BioLector software!

DO calibration parameters Lot No.2114211 (BioLector® Pro, filter module ID-228/-428)

Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
ϕ cal0	70.83	70.82	70.82	70.81	70.80	70.79	70.78
ϕ cal100	41.46	41.28	41.11	40.93	40.76	40.58	40.41

Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
ϕ cal0	70.78	70.77	70.76	70.75	70.74	70.74	70.73
ϕ cal100	40.23	40.06	39.88	39.71	39.53	39.36	39.18

Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
ϕ cal0	70.72	70.71	70.70	70.70	70.69	70.68	70.67
ϕ cal100	39.01	38.83	38.66	38.48	38.31	38.13	37.96

DO sensor properties

Dynamic range	0 - 100 % air saturation (a.s.)
Resolution	Up to 0.1 % O ₂ (software)
Accuracy	± 5% dissolved oxygen (batch calibration)
Drift at 0% oxygen	< 0.5% O ₂ per day (sampling interval of 6 min)
Response time (t ₉₀)	< 30 s
Temperature range	5 – 50°C
Sensor stability	Sensor material can be degraded by some microorganisms
Cross-sensitivity to	Organic solvents, such as acetone, toluene, chloroform or methylene chloride, Chlorine gas; high concentration of fluorescent molecules in the visible range can interfere (mCherry, tdTomato, dsRed, Nile red); complex media can cause a DO-shift
Basic material	Oxygen sensor RF-21160296 (at least stable for 7 days with CertiPUR-buffer) DO sensors are light-sensitive; please protect them from direct light!

DO calibration

Calibration	0.5 M Sulfite system (Two-point calibration with oxygen-free environment (sodium sulfite) and air-saturated environment)
Settings	BioLector protocol = LG1-RF-calibration, T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Microfluidic Flower Plate (MTP-MF32-BOH2)
Calibration device	Hardware ID: BL-09-000F-0032
Calibration phase offset	DO -360.44 (DO Ser. 4302-RD, gain 4)
Date of calibration	2021-10-22

Sterilization procedure

Sterilization	Beta irradiation (20 kGy)
BGS-certificate No	952124
Date of sterilization	2021-10-05

HEADQUARTERS EUROPE

m2p-labs GmbH
Arnold-Sommerfeld-Ring 2
52499 Baesweiler, Germany
Phone +49 -2401 805 330
Fax +49 -2401 805 33
info@m2p-labs.com

SUPPORT

EUROPE
Phone +49 - 2401 805 335
support@m2p-labs.com

N./S. AMERICAS
Phone +1 631 501 1878
supportUS@m2p-labs.com

ASIA PACIFIC
Phone +852 6092 6778
supportAsia@m2p-labs.com