

Please enter these calibration parameters and the Lot No. into the BioLector software!

**pH calibration parameters Lot No.2117101 and 2117107 (BioLector XT Microbioreactor, filter module ID-502)**

Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
φ min	64.82	64.73	64.65	64.57	64.49	64.40	64.32
φ max	14.16	14.16	14.16	14.17	14.17	14.17	14.18
dpH	0.56	0.56	0.56	0.56	0.56	0.56	0.56
pH <sub>0</sub>	6.17	6.16	6.16	6.15	6.14	6.13	6.13

Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
φ min	64.24	64.16	64.07	63.99	63.91	63.83	63.74
φ max	14.18	14.18	14.19	14.19	14.19	14.20	14.20
dpH	0.56	0.56	0.56	0.56	0.56	0.56	0.56
pH <sub>0</sub>	6.12	6.11	6.10	6.10	6.09	6.08	6.07

Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
φ min	63.66	63.58	63.50	63.41	63.33	63.25	63.17
φ max	14.20	14.21	14.21	14.22	14.22	14.22	14.23
dpH	0.56	0.56	0.57	0.57	0.57	0.57	0.57
pH <sub>0</sub>	6.07	6.06	6.05	6.04	6.04	6.03	6.02

**pH sensor properties**

Dynamic range	pH 4.20 - 7.75
Resolution	Up to 0.01 pH (software)
Accuracy	± 0.25 pH at pH 4.65 - 4.95 ; ± 0.1 pH at pH 4.95 - 6.90 ; ± 0.25 pH at pH 6.90 - 7.25 (batch calibration)
Response time (t90)	At 25 °C < 30 s
Drift at pH = 7	< 0.005 pH per day (sampling interval of 6 min)
Temperature range	5 °C to 50 °C
Compatibility	Aqueous solutions, ethanol, methanol (max. 5 % v/v)
Sensor stability	Sensor material can be degraded by some microorganisms
Cross-sensitivity	Reduced to ionic strength (salinity); high concentration of fluorescent molecules in the visible range can interfere (GFP, (e)YFP); complex media can cause a pH-shift (peptone, yeast extract)
Basic material	pH sensor HP8-1811-01_6 (at least stable for 7 days with CertiPUR-buffer)

**pH sensors are light-sensitive; please protect them from direct light!**

**pH calibration**

Buffer	CertiPUR Reference Material Buffer solutions Set (pH 3.00 ± 0.02 / pH 4.00 ± 0.02 / pH 9.00 ± 0.03 / pH 10.00 ± 0.03, 20 °C); 150 mM Citrat-Na-Phosphate buffer (16 solutions)
Settings	BioLector protocol = pH_DO_calibration_BOH1 ,T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Round Well Plate (MTP-R48-BOH1)
Calibration device	Hardware ID: 03166164 (BLXT Pilot 1)
Calibration phase offset	pH -2.45 (pH Ser. 3511, gain 7)
Date of calibration	2021-12-23

**Contact us**

- If you have any questions, contact Beckman Coulter Customer Support Center:
- Worldwide, find out in our website at: [www.beckman.de/support/technical](http://www.beckman.de/support/technical)
  - In the USA and Canada, call us at 1-800-369-0333
  - Outside the USA and Canada, contact your local Beckman Coulter representative

Please enter these calibration parameters and the Lot No. into the BioLection software!

**DO calibration parameters Lot No.2117101 and 2117107 (BioLector XT Microbioreactor, filter module ID-503)**

Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
A	-4257	-4189	-4121	-4054	-3986	-3918	-3850
B	33542	33005	32467	31929	31392	30854	30316
C	-30422	-29931	-29441	-28950	-28459	-27969	-27478

Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
A	-3782	-3714	-3646	-3578	-3511	-3443	-3375
B	29779	29241	28703	28166	27628	27090	26553
C	-26988	-26497	-26006	-25516	-25025	-24535	-24044

Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
A	-3307	-3239	-3171	-3103	-3035	-2968	-2900
B	26015	25477	24940	24402	23864	23327	22789
C	-23553	-23063	-22572	-22082	-21591	-21100	-20610

#### DO sensor properties

Dynamic range	0 - 100 % oxygen
Resolution	Up to 0.1 % O <sub>2</sub> (software)
Accuracy	± 5% dissolved oxygen (batch calibration)
Drift at 0% oxygen	< 0.5% O <sub>2</sub> per day (sampling interval of 6 min)
Response time (t <sub>90</sub> )	< 30 s
Temperature range	5 – 50°C
Sensor stability	sensor material can be degraded by some microorganisms
Cross-sensitivity to	Organic solvents, such as acetone, toluene, chloroform or methylene chloride, Chlorine gas; high concentration of fluorescent molecules in the visible range can interfere (mCherry, tdTomato, dsRed, Nile red); complex media can cause a DO-shift
Basic material	Oxygen sensor Pst3-HG-1921-01_2 (at least stable for 7 days with CertiPUR-buffer) <b>DO sensors are light-sensitive; please protect them from direct light!</b>

#### DO calibration

Calibration	0.5 M Sulfite system (Two-point calibration with oxygen-free environment (sodium sulfite) and air-saturated environment)
Settings	BioLector protocol = pH_DO_calibration_BOH1, T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Round Well Plate (MTP-R48-BOH1)
Calibration device	Hardware ID: 03166164 (BLXT Pilot 1)
Calibration phase offset	DO -360.83 (DO Ser. 4446, gain 7)
Date of calibration	2021-12-23

#### Sterilization procedure

Sterilization	Beta irradiation (20 kGy)
BGS-certificate No	979401
Date of sterilization	2021-12-14

#### Contact us

If you have any questions, contact Beckman Coulter Customer Support Center:

- Worldwide, find out in our website at: [www.beckman.de/support/technical](http://www.beckman.de/support/technical)
- In the USA and Canada, call us at 1-800-369-0333
- Outside the USA and Canada, contact your local Beckman Coulter representative