

Please enter these calibration parameters and the Lot No. into the BioLector software!

pH calibration parameters Lot No.2117101 and 2117107 (BioLector II/Pro Microbioreactor, filter module ID-202/402)

| Temperature | 20°C | 21°C | 22°C | 23°C | 24°C | 25°C | 26°C |
|-----------------|-------|-------|-------|-------|-------|-------|-------|
| φ min | 64.82 | 64.73 | 64.65 | 64.57 | 64.49 | 64.40 | 64.32 |
| φ max | 14.16 | 14.16 | 14.16 | 14.17 | 14.17 | 14.17 | 14.18 |
| dpH | 0.56 | 0.56 | 0.56 | 0.56 | 0.56 | 0.56 | 0.56 |
| pH ₀ | 6.17 | 6.16 | 6.16 | 6.15 | 6.14 | 6.13 | 6.13 |

| Temperature | 27°C | 28°C | 29°C | 30°C | 31°C | 32°C | 33°C |
|-----------------|-------|-------|-------|-------|-------|-------|-------|
| φ min | 64.24 | 64.16 | 64.07 | 63.99 | 63.91 | 63.83 | 63.74 |
| φ max | 14.18 | 14.18 | 14.19 | 14.19 | 14.19 | 14.20 | 14.20 |
| dpH | 0.56 | 0.56 | 0.56 | 0.56 | 0.56 | 0.56 | 0.56 |
| pH ₀ | 6.12 | 6.11 | 6.10 | 6.10 | 6.09 | 6.08 | 6.07 |

| Temperature | 34°C | 35°C | 36°C | 37°C | 38°C | 39°C | 40°C |
|-----------------|-------|-------|-------|-------|-------|-------|-------|
| φ min | 63.66 | 63.58 | 63.50 | 63.41 | 63.33 | 63.25 | 63.17 |
| φ max | 14.20 | 14.21 | 14.21 | 14.22 | 14.22 | 14.22 | 14.23 |
| dpH | 0.56 | 0.56 | 0.57 | 0.57 | 0.57 | 0.57 | 0.57 |
| pH ₀ | 6.07 | 6.06 | 6.05 | 6.04 | 6.04 | 6.03 | 6.02 |

pH sensor properties

| | |
|---------------------|---|
| Dynamic range | pH 4.20 - 7.75 |
| Resolution | Up to 0.01 pH (software) |
| Accuracy | ± 0.25 pH at pH 4.65 - 4.95 ; ± 0.1 pH at pH 4.95 - 6.90 ; ± 0.25 pH at pH 6.90 - 7.25 (batch calibration) |
| Response time (t90) | At 25 °C < 30 s |
| Drift at pH = 7 | < 0.005 pH per day (sampling interval of 6 min) |
| Temperature range | 5 °C to 50 °C |
| Compatibility | Aqueous solutions, ethanol, methanol (max. 5 % v/v) |
| Sensor stability | Sensor material can be degraded by some microorganisms |
| Cross-sensitivity | Reduced to ionic strength (salinity); high concentration of fluorescent molecules in the visible range can interfere (GFP, (e)YFP); complex media can cause a pH-shift (peptone, yeast extract) |
| Basic material | pH sensor HP8-1811-01_6 (at least stable for 7 days with CertiPUR-buffer) pH sensors are light-sensitive; please protect them from direct light! |

pH calibration

| | |
|--------------------------|---|
| Buffer | CertiPUR Reference Material Buffer solutions Set (pH 2.00 ± 0.01 / pH 3.00 ± 0.015 / pH 9.00 ± 0.01 / pH 10.00 ± 0.03, 20 °C); 150 mM Citrat-Na-Phosphate buffer (16 solutions) |
| Settings | BioLector protocol = pH_DO_calibration_BOH1 ,T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Round Well Plate (MTP-R48-BOH1) |
| Calibration device | Hardware ID: 03166164 (BLXT Pilot 1) |
| Calibration phase offset | pH -2.45 (pH Ser. 3511, gain 7) |
| Date of calibration | 2021-12-23 |

Contact us

If you have any questions, contact Beckman Coulter Customer Support Center:

- Worldwide, find out in our website at: www.beckman.de/support/technical
- In the USA and Canada, call us at 1-800-369-0333
- Outside the USA and Canada, contact your local Beckman Coulter representative

Please enter these calibration parameters and the Lot No. into the BioLector software!

DO calibration parameters Lot No.2117101 and 2117107 (BioLector II/Pro Microbioreactor, filter module ID-203/403)

| Temperature | 20°C | 21°C | 22°C | 23°C | 24°C | 25°C | 26°C |
|-------------|-------|-------|-------|-------|-------|-------|-------|
| φ cal0 | 73.11 | 73.06 | 73.01 | 72.96 | 72.91 | 72.86 | 72.81 |
| φ cal100 | 43.44 | 43.20 | 42.96 | 42.72 | 42.48 | 42.24 | 42.00 |

| Temperature | 27°C | 28°C | 29°C | 30°C | 31°C | 32°C | 33°C |
|-------------|-------|-------|-------|-------|-------|-------|-------|
| φ cal0 | 72.76 | 72.71 | 72.66 | 72.62 | 72.57 | 72.52 | 72.47 |
| φ cal100 | 41.76 | 41.52 | 41.28 | 41.05 | 40.81 | 40.57 | 40.33 |

| Temperature | 34°C | 35°C | 36°C | 37°C | 38°C | 39°C | 40°C |
|-------------|-------|-------|-------|-------|-------|-------|-------|
| φ cal0 | 72.42 | 72.37 | 72.32 | 72.27 | 72.22 | 72.17 | 72.13 |
| φ cal100 | 40.09 | 39.85 | 39.61 | 39.37 | 39.13 | 38.89 | 38.65 |

DO sensor properties

| | |
|----------------------------------|---|
| Dynamic range | 0 - 100 % air saturation (a.s.) |
| Resolution | Up to 0.1 % O ₂ (software) |
| Accuracy | ± 5% dissolved oxygen (batch calibration) |
| Drift at 0% oxygen | < 0.5% O ₂ per day (sampling interval of 6 min) |
| Response time (t ₉₀) | < 30 s |
| Temperature range | 5 – 50°C |
| Sensor stability | Sensor material can be degraded by some microorganisms |
| Cross-sensitivity to | Organic solvents, such as acetone, toluene, chloroform or methylene chloride, Chlorine gas; high concentration of fluorescent molecules in the visible range can interfere (mCherry, tdTomato, dsRed, Nile red); complex media can cause a DO-shift |
| Basic material | Oxygen sensor Pst3-HG-1921-01_2 (at least stable for 7 days with CertiPUR-buffer) DO sensors are light-sensitive; please protect them from direct light! |

DO calibration

| | |
|--------------------------|--|
| Calibration | 0.5 M Sulfite system (Two-point calibration with oxygen-free environment (sodium sulfite) and air-saturated environment) |
| Settings | BioLector protocol = pH_DO_calibration_BOH1 , T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Round Well Plate (MTP-R48-BOH1) |
| Calibration device | Hardware ID: 03166164 (BLXT Pilot 1) |
| Calibration phase offset | DO -360.83 (DO Ser. 4446, gain 7) |
| Date of calibration | 2021-12-23 |

Sterilization procedure

| | |
|-----------------------|---------------------------|
| Sterilization | Beta irradiation (20 kGy) |
| BGS-certificate No | 979401 |
| Date of sterilization | 2021-12-14 |

Contact us

If you have any questions, contact Beckman Coulter Customer Support Center:

- Worldwide, find out in our website at: www.beckman.de/support/technical
- In the USA and Canada, call us at 1-800-369-0333
- Outside the USA and Canada, contact your local Beckman Coulter representative