

Please enter these calibration parameters and the Lot No. into the BioLecton software!

**pH calibration parameters Lot No.2202102 (BioLector XT Microbioreactor, filter module ID-502)**

Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
φ min	64.98	64.89	64.80	64.71	64.62	64.54	64.45
φ max	12.86	12.86	12.86	12.86	12.87	12.87	12.87
dpH	0.55	0.55	0.55	0.55	0.55	0.55	0.55
pH <sub>0</sub>	6.21	6.21	6.20	6.19	6.18	6.17	6.17

Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
φ min	64.36	64.27	64.19	64.10	64.01	63.92	63.84
φ max	12.87	12.87	12.88	12.88	12.88	12.88	12.89
dpH	0.55	0.55	0.55	0.55	0.55	0.56	0.56
pH <sub>0</sub>	6.16	6.15	6.14	6.14	6.13	6.12	6.11

Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
φ min	63.75	63.66	63.57	63.49	63.40	63.31	63.22
φ max	12.89	12.89	12.89	12.90	12.90	12.90	12.90
dpH	0.56	0.56	0.56	0.56	0.56	0.56	0.56
pH <sub>0</sub>	6.11	6.10	6.09	6.08	6.08	6.07	6.06

**pH sensor properties**

Dynamic range	pH 4.20 - 7.80
Resolution	Up to 0.01 pH (software)
Accuracy	± 0.25 pH at pH 4.70 - 5.00 ; ± 0.1 pH at pH 5.00 - 7.00 ; ± 0.25 pH at pH 7.00 - 7.30 (batch calibration)
Response time (t90)	At 25 °C < 30 s
Drift at pH = 7	< 0.005 pH per day (sampling interval of 6 min)
Temperature range	5 °C to 50 °C
Compatibility	Aqueous solutions, ethanol, methanol (max. 5 % v/v)
Sensor stability	Sensor material can be degraded by some microorganisms
Cross-sensitivity	Reduced to ionic strength (salinity); high concentration of fluorescent molecules in the visible range can interfere (GFP, (e)YFP); complex media can cause a pH-shift (peptone, yeast extract)
Basic material	pH sensor HP8-1811-01_6 (at least stable for 7 days with CertiPUR-buffer)

**pH sensors are light-sensitive; please protect them from direct light!**

**pH calibration**

Buffer	CertiPUR Reference Material Buffer solutions Set (pH 3.00 ± 0.02 / pH 4.00 ± 0.02 / pH 9.00 ± 0.03 / pH 10.00 ± 0.03, 20 °C); 150 mM Citrat-Na-Phosphate buffer (16 solutions)
Settings	BioLector protocol = pH_DO_calibration_BOH1 ,T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Flower Plate (MTP-48-BOH1)
Calibration device	Hardware ID: 03166164 (BLXT Pilot 1)
Calibration phase offset	pH -2.07 (pH Ser. 3567, gain 7)
Date of calibration	2022-03-02

**Contact us**

- If you have any questions, contact Beckman Coulter Customer Support Center:
- Worldwide, find out in our website at: [www.beckman.de/support/technical](http://www.beckman.de/support/technical)
  - In the USA and Canada, call us at 1-800-369-0333
  - Outside the USA and Canada, contact your local Beckman Coulter representative

Please enter these calibration parameters and the Lot No. into the BioLecton software!

**DO calibration parameters Lot No.2202102 (BioLector XT Microbioreactor, filter module ID-503)**

Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
A	-4345	-4271	-4198	-4124	-4051	-3977	-3904
B	34244	33662	33080	32498	31916	31334	30752
C	-31069	-30538	-30007	-29475	-28944	-28413	-27881

Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
A	-3831	-3757	-3684	-3610	-3537	-3463	-3390
B	30170	29588	29006	28424	27842	27261	26679
C	-27350	-26819	-26287	-25756	-25225	-24693	-24162

Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
A	-3317	-3243	-3170	-3096	-3023	-2950	-2876
B	26097	25515	24933	24351	23769	23187	22605
C	-23631	-23100	-22568	-22037	-21506	-20974	-20443

#### DO sensor properties

Dynamic range	0 - 100 % oxygen
Resolution	Up to 0.1 % O <sub>2</sub> (software)
Accuracy	± 5% dissolved oxygen (batch calibration)
Drift at 0% oxygen	< 0.5% O <sub>2</sub> per day (sampling interval of 6 min)
Response time (t <sub>90</sub> )	< 30 s
Temperature range	5 – 50°C
Sensor stability	sensor material can be degraded by some microorganisms
Cross-sensitivity to	Organic solvents, such as acetone, toluene, chloroform or methylene chloride, Chlorine gas; high concentration of fluorescent molecules in the visible range can interfere (mCherry, tdTomato, dsRed, Nile red); complex media can cause a DO-shift
Basic material	Oxygen sensor PSt3-HG-1921-01_4 (at least stable for 7 days with CertiPUR-buffer) <b>DO sensors are light-sensitive; please protect them from direct light!</b>

#### DO calibration

Calibration	Three-point calibration at an oxygen-free environment (1.0 M sulfite system), an air-saturated environment (21% oxygen) and a pure (100%) oxygen environment (latter two with OC buffer)
Settings	BioLector protocol = pH_DO_calibration_BOH1, T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Flower Plate (MTP-48-BOH1)
Calibration device	Hardware ID: 03166164 (BLXT Pilot 1)
Calibration phase offset	DO -360.98 (DO Ser. 4446, gain 7)
Date of calibration	2022-03-02

#### Sterilization procedure

Sterilization	Beta irradiation (20 kGy)
BGS-certificate No	1001695
Date of sterilization	2022-02-16

#### Contact us

- If you have any questions, contact Beckman Coulter Customer Support Center:
- Worldwide, find out in our website at: [www.beckman.de/support/technical](http://www.beckman.de/support/technical)
  - In the USA and Canada, call us at 1-800-369-0333
  - Outside the USA and Canada, contact your local Beckman Coulter representative