

Please enter these **calibration parameters** and the **Lot No.** into the BioLector software!

pH calibration parameters Lot No.2204311 (BioLector Pro Microbioreactor, filter module ID-424)

Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
ϕ min	68.18	68.29	68.41	68.52	68.64	68.75	68.87
ϕ max	13.17	13.21	13.24	13.27	13.31	13.34	13.37
dpH	-0.41	-0.41	-0.41	-0.41	-0.41	-0.41	-0.41
pH ₀	5.51	5.50	5.50	5.49	5.48	5.48	5.47

Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
ϕ min	68.98	69.10	69.21	69.33	69.44	69.56	69.67
ϕ max	13.40	13.44	13.47	13.50	13.54	13.57	13.60
dpH	-0.41	-0.41	-0.41	-0.40	-0.40	-0.40	-0.40
pH ₀	5.46	5.46	5.45	5.44	5.44	5.43	5.42

Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
ϕ min	69.79	69.90	70.02	70.13	70.25	70.36	70.48
ϕ max	13.63	13.67	13.70	13.73	13.77	13.80	13.83
dpH	-0.40	-0.40	-0.40	-0.40	-0.40	-0.40	-0.40
pH ₀	5.42	5.41	5.40	5.40	5.39	5.38	5.38

pH sensor properties

Dynamic range	pH 3.95 - 6.65
Resolution	Up to 0.01 pH (software)
Accuracy	± 0.25 pH at pH 4.25-4.45; ± 0.1 pH at pH 4.45-6.15; ± 0.25 pH at pH 6.15-6.35 batch calibration
Response time (t90)	At 25 °C < 30 s
Drift at pH = 7	< 0.005 pH per day (sampling interval of 6 min)
Temperature range	5 °C to 50 °C
Compatibility	Aqueous solutions, ethanol, methanol (max. 5 % v/v)
Sensor stability	Sensor material can be degraded by some microorganisms
Cross-sensitivity	Reduced to ionic strength (salinity); high concentration of fluorescent molecules in the visible range can interfere (GFP, (e)YFP); complex media can cause a pH-shift (peptone, yeast extract)
Basic material	pH sensor pH51-211650283 (at least stable for 7 days with CertiPUR-buffer) pH sensors are light-sensitive; please protect them from direct light!

pH calibration

Buffer	CertiPUR Reference Material Buffer solutions Set (pH 2.00 ± 0.02 / pH 3.00 ± 0.02 / pH 7.00 ± 0.02 / pH 8.00 ± 0.03, 20 °C); 150 mM Citrat-Na-Phosphate buffer (16 solutions)
Settings	BioLector protocol = pH51-RF-calibration, T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Microfluidic Flower Plate (MTP-MF32-BOH3)
Calibration device	Hardware ID: BL-09-000F-0032
Calibration phase offset	pH -360.10 (pH Ser. 3288, gain 6)
Date of calibration	2022-04-13

Contact us

If you have any questions, contact Beckman Coulter Customer Support Center:

- Worldwide, find out in our website at: www.beckman.de/support/technical
- In the USA and Canada, call us at 1-800-369-0333
- Outside the USA and Canada, contact your local Beckman Coulter representative

Please enter these **calibration parameters** and the **Lot No.** into the BioLector software!

DO calibration parameters Lot No.2204311 (BioLector Pro Microbioreactor, filter module ID-228/-428)

Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
ϕ cal0	70.37	70.36	70.34	70.33	70.31	70.30	70.28
ϕ cal100	41.01	40.85	40.69	40.53	40.37	40.21	40.05

Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
ϕ cal0	70.27	70.25	70.24	70.22	70.20	70.19	70.17
ϕ cal100	39.89	39.73	39.57	39.41	39.25	39.09	38.93

Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
ϕ cal0	70.16	70.14	70.13	70.11	70.10	70.08	70.07
ϕ cal100	38.77	38.61	38.44	38.28	38.12	37.96	37.80

DO sensor properties

Dynamic range	0 - 100 % air saturation (a.s.)
Resolution	Up to 0.1 % O ₂ (software)
Accuracy	± 5% dissolved oxygen (batch calibration)
Drift at 0% oxygen	< 0.5% O ₂ per day (sampling interval of 6 min)
Response time (t ₉₀)	< 30 s
Temperature range	5 – 50°C
Sensor stability	Sensor material can be degraded by some microorganisms
Cross-sensitivity to	Organic solvents, such as acetone, toluene, chloroform or methylene chloride, Chlorine gas; high concentration of fluorescent molecules in the visible range can interfere (mCherry, tdTomato, dsRed, Nile red); complex media can cause a DO-shift
Basic material	Oxygen sensor RF-213550643 (at least stable for 7 days with CertiPUR-buffer) DO sensors are light-sensitive; please protect them from direct light!

DO calibration

Calibration	1.0 M Sulfite system (Two-point calibration with oxygen-free environment (sodium sulfite) and air-saturated environment)
Settings	BioLector protocol = pH51-RF-calibration, T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Microfluidic Flower Plate (MTP-MF32-BOH3)
Calibration device	Hardware ID: BL-09-000F-0032
Calibration phase offset	DO -360.44 (DO Ser. 4302-RD, gain 4)
Date of calibration	2022-04-13

Sterilization procedure

Sterilization	Beta irradiation (20 kGy)
BGS-certificate No	1020859
Date of sterilization	2022-04-01

Contact us

If you have any questions, contact Beckman Coulter Customer Support Center:

- Worldwide, find out in our website at: www.beckman.de/support/technical
- In the USA and Canada, call us at 1-800-369-0333
- Outside the USA and Canada, contact your local Beckman Coulter representative