

Please enter these calibration parameters and the Lot No. into the BioLecton software!

pH calibration parameters Lot No.2205101 (BioLector XT Microbioreactor, filter module ID-502)

Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
φ min	64.48	64.41	64.34	64.27	64.20	64.14	64.07
φ max	13.35	13.34	13.34	13.34	13.34	13.33	13.33
d pH	0.56	0.56	0.56	0.56	0.56	0.56	0.56
pH ₀	6.29	6.28	6.27	6.26	6.26	6.25	6.24

Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
φ min	64.00	63.93	63.86	63.79	63.72	63.65	63.59
φ max	13.33	13.32	13.32	13.32	13.31	13.31	13.31
d pH	0.56	0.56	0.56	0.56	0.56	0.56	0.56
pH ₀	6.23	6.22	6.21	6.20	6.20	6.19	6.18

Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
φ min	63.52	63.45	63.38	63.31	63.24	63.17	63.10
φ max	13.30	13.30	13.30	13.30	13.29	13.29	13.29
d pH	0.56	0.56	0.56	0.56	0.56	0.56	0.56
pH ₀	6.17	6.16	6.15	6.14	6.13	6.13	6.12

pH sensor properties

Dynamic range	pH 4.25 - 7.85
Resolution	Up to 0.01 pH (software)
Accuracy	± 0.25 pH at pH 4.75 - 5.05 ; ± 0.1 pH at pH 5.05 - 7.05 ; ± 0.25 pH at pH 7.05 - 7.35 (batch calibration)
Response time (t ₉₀)	At 25 °C < 30 s
Drift at pH = 7	< 0.005 pH per day (sampling interval of 6 min)
Temperature range	5 °C to 50 °C
Compatibility	Aqueous solutions, ethanol, methanol (max. 5 % v/v)
Sensor stability	Sensor material can be degraded by some microorganisms
Cross-sensitivity	Reduced to ionic strength (salinity); high concentration of fluorescent molecules in the visible range can interfere (GFP, (e)YFP); complex media can cause a pH-shift (peptone, yeast extract)
Basic material	pH sensor HP8-1811-01_6 (at least stable for 7 days with CertiPUR-buffer) pH sensors are light-sensitive; please protect them from direct light!

pH calibration

Buffer	CertiPUR Reference Material Buffer solutions Set (pH 3.00 ± 0.02 / pH 4.00 ± 0.02 / pH 9.00 ± 0.03 / pH 10.00 ± 0.03, 20 °C); 150 mM Citrat-Na-Phosphate buffer (16 solutions)
Settings	BioLector protocol = pH_DO_calibration_BOH1, T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Flower Plate (MTP-48-BOH1)
Calibration device	Hardware ID: 03166164 (BLXT Pilot 1)
Calibration phase offset	pH -2.06 (pH Ser. 3567, gain 7)
Date of calibration	2022-05-11

Contact us

If you have any questions, contact Beckman Coulter Customer Support Center:

- Worldwide, find out in our website at: www.beckman.de/support/technical
- In the USA and Canada, call us at 1-800-369-0333
- Outside the USA and Canada, contact your local Beckman Coulter representative

Please enter these **calibration parameters** and the **Lot No.** into the BioLector software!

DO calibration parameters Lot No.2205101 (BioLector XT Microbioreactor, filter module ID-503)

Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
A	-4553	-4472	-4390	-4308	-4226	-4144	-4062
B	35919	35268	34617	33967	33316	32666	32015
C	-32618	-32023	-31428	-30833	-30238	-29643	-29048

Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
A	-3980	-3898	-3816	-3734	-3652	-3570	-3488
B	31365	30714	30063	29413	28762	28112	27461
C	-28453	-27857	-27262	-26667	-26072	-25477	-24882

Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
A	-3406	-3324	-3242	-3160	-3078	-2996	-2915
B	26811	26160	25509	24859	24208	23558	22907
C	-24287	-23692	-23096	-22501	-21906	-21311	-20716

DO sensor properties

Dynamic range	0 - 100 % oxygen
Resolution	Up to 0.1 % O ₂ (software)
Accuracy	± 5% dissolved oxygen (batch calibration)
Drift at 0% oxygen	< 0.5% O ₂ per day (sampling interval of 6 min)
Response time (t ₉₀)	< 30 s
Temperature range	5 – 50°C
Sensor stability	sensor material can be degraded by some microorganisms
Cross-sensitivity to	Organic solvents, such as acetone, toluene, chloroform or methylene chloride, Chlorine gas; high concentration of fluorescent molecules in the visible range can interfere (mCherry, tdTomato, dsRed, Nile red); complex media can cause a DO-shift
Basic material	Oxygen sensor PSt3-HG-1921-01_4 (at least stable for 7 days with CertiPUR-buffer)

DO sensors are light-sensitive; please protect them from direct light!

DO calibration

Calibration	Three-point calibration at an oxygen-free environment (1.0 M sulfite system), an air-saturated environment (21% oxygen) and a pure (100%) oxygen environment (latter two with QC buffer)
Settings	BioLector protocol = pH_DO_calibration_BOH1 , T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Flower Plate (MTP-48-BOH1)
Calibration device	Hardware ID: 03166164 (BLXT Pilot 1)
Calibration phase offset	DO -360.98 (DO Ser. 4446, gain 7)
Date of calibration	2022-05-11

Sterilization procedure

Sterilization	Beta irradiation (20 kGy)
BGS-certificate No	1029868
Date of sterilization	2022-04-26

Contact us

If you have any questions, contact Beckman Coulter Customer Support Center:

- Worldwide, find out in our website at: www.beckman.de/support/technical
- In the USA and Canada, call us at 1-800-369-0333
- Outside the USA and Canada, contact your local Beckman Coulter representative