

Please enter these calibration parameters and the Lot No. into the BioLector software!

pH calibration parameters Lot No.2212201+2212207 (BioLector II/Pro Microbioreactor, filter module ID-221/421)

| Temperature | 20°C | 21°C | 22°C | 23°C | 24°C | 25°C | 26°C |
|-----------------|-------|-------|-------|-------|-------|-------|-------|
| φ min | 72.99 | 72.87 | 72.76 | 72.65 | 72.53 | 72.42 | 72.31 |
| φ max | 24.86 | 24.76 | 24.66 | 24.55 | 24.45 | 24.35 | 24.25 |
| dpH | 0.66 | 0.66 | 0.66 | 0.67 | 0.67 | 0.67 | 0.67 |
| pH ₀ | 5.88 | 5.87 | 5.86 | 5.85 | 5.84 | 5.83 | 5.82 |

| Temperature | 27°C | 28°C | 29°C | 30°C | 31°C | 32°C | 33°C |
|-----------------|-------|-------|-------|-------|-------|-------|-------|
| φ min | 72.19 | 72.08 | 71.97 | 71.85 | 71.74 | 71.62 | 71.51 |
| φ max | 24.15 | 24.05 | 23.95 | 23.84 | 23.74 | 23.64 | 23.54 |
| dpH | 0.67 | 0.67 | 0.67 | 0.67 | 0.67 | 0.67 | 0.67 |
| pH ₀ | 5.82 | 5.81 | 5.80 | 5.79 | 5.78 | 5.77 | 5.77 |

| Temperature | 34°C | 35°C | 36°C | 37°C | 38°C | 39°C | 40°C |
|-----------------|-------|-------|-------|-------|-------|-------|-------|
| φ min | 71.40 | 71.28 | 71.17 | 71.06 | 70.94 | 70.83 | 70.71 |
| φ max | 23.44 | 23.34 | 23.24 | 23.14 | 23.03 | 22.93 | 22.83 |
| dpH | 0.67 | 0.67 | 0.67 | 0.67 | 0.67 | 0.67 | 0.67 |
| pH ₀ | 5.76 | 5.75 | 5.74 | 5.73 | 5.72 | 5.71 | 5.71 |

pH sensor properties

| | |
|---------------------|---|
| Dynamic range | pH 3.55 - 7.70 |
| Resolution | Up to 0.01 pH (software) |
| Accuracy | ± 0.25 pH at pH 4.20 - 4.65 ; ± 0.1 pH at pH 4.65 - 6.60 ; ± 0.25 pH at pH 6.60 - 7.05 (batch calibration) |
| Response time (t90) | At 25 °C < 30 s |
| Drift at pH = 7 | < 0.005 pH per day (sampling interval of 6 min) |
| Temperature range | 5 °C to 50 °C |
| Compatibility | Aqueous solutions, ethanol, methanol (max. 5 % v/v) |
| Sensor stability | Sensor material can be degraded by some microorganisms |
| Cross-sensitivity | Reduced to ionic strength (salinity); high concentration of fluorescent molecules in the visible range can interfere (GFP, (e)YFP); complex media can cause a pH-shift (peptone, yeast extract) |
| Basic material | pH sensor LG1-2141-01 (at least stable for 7 days with CertiPUR-buffer) pH sensors are light-sensitive; please protect them from direct light! |

pH calibration

| | |
|--------------------------|--|
| Buffer | CertiPUR Reference Material Buffer solutions Set (pH 2.00 ± 0.02 / pH 3.00 ± 0.02 / pH 9.00 ± 0.03 / pH 10.00 ± 0.03, 20 °C); 150 mM Citrat-Na-Phosphate buffer (16 solutions) |
| Settings | BioLector protocol = pH_DO_calibration_BOH2 ,T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Round Well Plate (MTP-R48-BOH2) |
| Calibration device | Hardware ID: 03166164 (BLXT Pilot 1) |
| Calibration phase offset | pH -360.63 (pH Ser. 3513, gain 8) |
| Date of calibration | 2022-09-23 |

Contact us

If you have any questions, contact Beckman Coulter Customer Support Center:

- Worldwide, find out in our website at: www.beckman.de/support/technical
- In the USA and Canada, call us at 1-800-369-0333
- Outside the USA and Canada, contact your local Beckman Coulter representative

Please enter these calibration parameters and the Lot No. into the BioLector software!

DO calibration parameters Lot No.2212201+2212207 (BioLector II/Pro Microbioreactor, filter module ID-228/428)

| Temperature | 20°C | 21°C | 22°C | 23°C | 24°C | 25°C | 26°C |
|-------------|-------|-------|-------|-------|-------|-------|-------|
| φ cal0 | 70.42 | 70.40 | 70.38 | 70.36 | 70.33 | 70.31 | 70.29 |
| φ cal100 | 41.24 | 41.03 | 40.82 | 40.61 | 40.40 | 40.19 | 39.98 |

| Temperature | 27°C | 28°C | 29°C | 30°C | 31°C | 32°C | 33°C |
|-------------|-------|-------|-------|-------|-------|-------|-------|
| φ cal0 | 70.27 | 70.25 | 70.23 | 70.21 | 70.19 | 70.16 | 70.14 |
| φ cal100 | 39.77 | 39.56 | 39.35 | 39.14 | 38.93 | 38.72 | 38.51 |

| Temperature | 34°C | 35°C | 36°C | 37°C | 38°C | 39°C | 40°C |
|-------------|-------|-------|-------|-------|-------|-------|-------|
| φ cal0 | 70.12 | 70.10 | 70.08 | 70.06 | 70.04 | 70.01 | 69.99 |
| φ cal100 | 38.30 | 38.09 | 37.89 | 37.68 | 37.47 | 37.26 | 37.05 |

DO sensor properties

| | |
|----------------------------------|---|
| Dynamic range | 0 - 100 % air saturation (a.s.) |
| Resolution | Up to 0.1 % O ₂ (software) |
| Accuracy | ± 5% dissolved oxygen (batch calibration) |
| Drift at 0% oxygen | < 0.5% O ₂ per day (sampling interval of 6 min) |
| Response time (t ₉₀) | < 30 s |
| Temperature range | 5 – 50°C |
| Sensor stability | Sensor material can be degraded by some microorganisms |
| Cross-sensitivity to | Organic solvents, such as acetone, toluene, chloroform or methylene chloride, Chlorine gas; high concentration of fluorescent molecules in the visible range can interfere (mCherry, tdTomato, dsRed, Nile red); complex media can cause a DO-shift |
| Basic material | Oxygen sensor RF- 221155398 (at least stable for 7 days with CertiPUR-buffer) DO sensors are light-sensitive; please protect them from direct light! |

DO calibration

| | |
|--------------------------|---|
| Calibration | Two-point calibration at an oxygen-free environment (1.0 M sulfite system) and an air-saturated environment (21% oxygen with QC buffer) |
| Settings | BioLector protocol = pH_DO_calibration_BOH2, T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Round Well Plate (MTP-R48-BOH2) |
| Calibration device | Hardware ID: 03166164 (BLXT Pilot 1) |
| Calibration phase offset | DO -360.94 (DO Ser. 4452, gain 4) |
| Date of calibration | 2022-09-23 |

Sterilization procedure

| | |
|-----------------------|---------------------------|
| Sterilization | Beta irradiation (20 kGy) |
| BGS-certificate No | 1078918 |
| Date of sterilization | 2022-09-07 |

Contact us

If you have any questions, contact Beckman Coulter Customer Support Center:

- Worldwide, find out in our website at: www.beckman.de/support/technical
- In the USA and Canada, call us at 1-800-369-0333
- Outside the USA and Canada, contact your local Beckman Coulter representative