

Please enter these **calibration parameters** and the **Lot No.** into the BioLecton software!

pH calibration parameters Lot No.2215201 (BioLector XT Microbioreactor, filter module ID-521)

Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
φ min	69.91	69.82	69.73	69.65	69.56	69.48	69.39
φ max	21.78	21.72	21.66	21.60	21.54	21.47	21.41
dpH	0.72	0.72	0.72	0.72	0.72	0.71	0.71
pH ₀	5.87	5.86	5.84	5.83	5.81	5.80	5.78

Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
φ min	69.30	69.22	69.13	69.05	68.96	68.87	68.79
φ max	21.35	21.29	21.23	21.16	21.10	21.04	20.98
dpH	0.71	0.71	0.71	0.71	0.71	0.71	0.71
pH ₀	5.77	5.76	5.74	5.73	5.71	5.70	5.69

Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
φ min	68.70	68.62	68.53	68.44	68.36	68.27	68.19
φ max	20.92	20.85	20.79	20.73	20.67	20.61	20.54
dpH	0.71	0.70	0.70	0.70	0.70	0.70	0.70
pH ₀	5.67	5.66	5.64	5.63	5.61	5.60	5.59

pH sensor properties

Dynamic range	pH 3.40 - 7.70
Resolution	Up to 0.01 pH (software)
Accuracy	± 0.25 pH at pH 4.15 - 4.65 ; ± 0.1 pH at pH 4.65 - 6.45 ; ± 0.25 pH at pH 6.45 - 6.95 (batch calibration)
Response time (t90)	At 25 °C < 30 s
Drift at pH = 7	< 0.005 pH per day (sampling interval of 6 min)
Temperature range	5 °C to 50 °C
Compatibility	Aqueous solutions, ethanol, methanol (max. 5 % v/v)
Sensor stability	Sensor material can be degraded by some microorganisms
Cross-sensitivity	Reduced to ionic strength (salinity); high concentration of fluorescent molecules in the visible range can interfere (GFP, (e)YFP); complex media can cause a pH-shift (peptone, yeast extract)
Basic material	pH sensor LG1-2212-01 (at least stable for 7 days with CertiPUR-buffer)

pH sensors are light-sensitive; please protect them from direct light!

pH calibration

Buffer	CertiPUR Reference Material Buffer solutions Set (pH 2.00 ± 0.02 / pH 3.00 ± 0.02 / pH 9.00 ± 0.03 / pH 10.00 ± 0.03, 20 °C); 150 mM Citrat-Na-Phosphate buffer (16 solutions)
Settings	BioLector protocol = pH_DO_calibration_BOH2 ,T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Flower Plate (MTP-48-BOH2)
Calibration device	Hardware ID: 03624952 (BLXT 0071)
Calibration phase offset	pH -360.44 (pH Ser. 3801, gain 8)
Date of calibration	2022-12-14

Contact us

- If you have any questions, contact Beckman Coulter Customer Support Center:
- Worldwide, find out in our website at: www.beckman.de/support/technical
 - In the USA and Canada, call us at 1-800-369-0333
 - Outside the USA and Canada, contact your local Beckman Coulter representative

Please enter these **calibration parameters** and the **Lot No.** into the BioLector software!

DO calibration parameters Lot No.2215201 (BioLector XT Microbioreactor, filter module ID-528)

Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
A	-1945	-1904	-1862	-1820	-1778	-1736	-1695
B	14994	14665	14337	14009	13680	13352	13024
C	-13250	-12953	-12656	-12359	-12062	-11765	-11468

Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
A	-1653	-1611	-1569	-1528	-1486	-1444	-1402
B	12695	12367	12038	11710	11382	11053	10725
C	-11171	-10874	-10577	-10280	-9983	-9686	-9389

Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
A	-1360	-1319	-1277	-1235	-1193	-1151	-1110
B	10397	10068	9740	9411	9083	8755	8426
C	-9092	-8795	-8498	-8201	-7904	-7607	-7310

DO sensor properties

Dynamic range	0 - 100 % oxygen
Resolution	Up to 0.1 % O ₂ (software)
Accuracy	± 5% dissolved oxygen (batch calibration)
Drift at 0% oxygen	< 0.5% O ₂ per day (sampling interval of 6 min)
Response time (t ₉₀)	< 30 s
Temperature range	5 – 50°C
Sensor stability	sensor material can be degraded by some microorganisms
Cross-sensitivity to	Organic solvents, such as acetone, toluene, chloroform or methylene chloride, Chlorine gas; high concentration of fluorescent molecules in the visible range can interfere (mCherry, tdTomato, dsRed, Nile red); complex media can cause a DO-shift
Basic material	Oxygen sensor RF-222756999 (at least stable for 7 days with CertiPUR-buffer) DO sensors are light-sensitive; please protect them from direct light!

DO calibration

Calibration	Three-point calibration at an oxygen-free environment (1.0 M sulfite system), an air-saturated environment (21% oxygen) and a pure (100%) oxygen environment (latter two with QC buffer)
Settings	BioLector protocol = pH_DO_calibration_BOH2 ,T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Flower Plate (MTP-48-BOH2)
Calibration device	Hardware ID: 03624952 (BLXT 0071)
Calibration phase offset	DO -360.78 (DO Ser. 4673, gain 4)
Date of calibration	2022-12-14

Sterilization procedure

Sterilization	Beta irradiation (20 kGy)
Steris Process Run ID	2324-3130
Date of sterilization	2022-12-06

Contact us

If you have any questions, contact Beckman Coulter Customer Support Center:

- Worldwide, find out in our website at: www.beckman.de/support/technical
- In the USA and Canada, call us at 1-800-369-0333
- Outside the USA and Canada, contact your local Beckman Coulter representative