

Please enter these calibration parameters and the Lot No. into the BioLecton software!

**pH calibration parameters Lot No.2304201 and 2304207 (BioLector® II/Pro, filter module ID-221/-421)**

Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
φ min	69.34	69.27	69.20	69.13	69.06	68.99	68.92
φ max	24.12	24.06	24.00	23.94	23.89	23.83	23.77
dpH	0.73	0.73	0.73	0.72	0.72	0.72	0.72
pH <sub>0</sub>	6.01	6.00	5.99	5.98	5.97	5.96	5.95

Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
φ min	68.86	68.79	68.72	68.65	68.58	68.51	68.44
φ max	23.71	23.65	23.59	23.53	23.47	23.41	23.36
dpH	0.72	0.72	0.72	0.72	0.72	0.72	0.72
pH <sub>0</sub>	5.94	5.93	5.92	5.91	5.90	5.89	5.88

Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
φ min	68.37	68.30	68.23	68.16	68.09	68.02	67.95
φ max	23.30	23.24	23.18	23.12	23.06	23.00	22.94
dpH	0.72	0.72	0.72	0.72	0.72	0.72	0.72
pH <sub>0</sub>	5.87	5.86	5.85	5.84	5.82	5.81	5.80

**pH sensor properties**

Dynamic range	pH 3.60 - 7.95
Resolution	Up to 0.01 pH (software)
Accuracy	± 0.25 pH at pH 4.35-4.90; ± 0.1 pH at pH 4.90-6.65; ± 0.25 pH at pH 6.65-7.20 batch calibration
Response time (t <sub>90</sub> )	At 25 °C < 30 s
Drift at pH = 7	< 0.005 pH per day (sampling interval of 6 min)
Temperature range	5 °C to 50 °C
Compatibility	Aqueous solutions, ethanol, methanol (max. 5 % v/v)
Sensor stability	Sensor material can be degraded by some microorganisms
Cross-sensitivity	Reduced to ionic strength (salinity); high concentration of fluorescent molecules in the visible range can interfere (GFP, (e)YFP); complex media can cause a pH-shift (peptone, yeast extract)
Basic material	pH sensor LG1-2239-01 (at least stable for 7 days with CertiPUR-buffer) <b>pH sensors are light-sensitive; please protect them from direct light!</b>

**pH calibration**

Buffer	CertiPUR Reference Material Buffer solutions Set (pH 2.00 ± 0.02 / pH 3.00 ± 0.02 / pH 9.00 ± 0.03 / pH 10.00 ± 0.03, 20 °C); 150 mM Citrat-Na-Phosphate buffer (16 solutions)
Settings	BioLector protocol = LG1-RF-calibration, T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Flower Plate (MTP-48-BOH2)
Calibration device	Hardware ID: BL-09-000F-0032
Calibration phase offset	pH -360.31 (pH Ser. 3305, gain 8)
Date of calibration	2023-03-17

**Contact us**

If you have any questions, contact Beckman Coulter Customer Support Center:

- Worldwide, find out in our website at: [www.beckman.de/support/technical](http://www.beckman.de/support/technical)
- In the USA and Canada, call us at 1-800-369-0333
- Outside the USA and Canada, contact your local Beckman Coulter representative

Please enter these calibration parameters and the Lot No. into the BioLector software!

**DO calibration parameters Lot No.2304201 and 2304207 (BioLector® II/Pro, filter module ID-228/-428)**

Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
φ cal0	70.87	70.85	70.82	70.80	70.78	70.75	70.73
φ cal100	41.70	41.52	41.35	41.17	40.99	40.82	40.64

Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
φ cal0	70.70	70.68	70.66	70.63	70.61	70.59	70.56
φ cal100	40.47	40.29	40.11	39.94	39.76	39.59	39.41

Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
φ cal0	70.54	70.52	70.49	70.47	70.44	70.42	70.40
φ cal100	39.23	39.06	38.88	38.70	38.53	38.35	38.18

**DO sensor properties**

Dynamic range	0 - 100 % air saturation (a.s.)
Resolution	Up to 0.1 % O <sub>2</sub> (software)
Accuracy	± 5% dissolved oxygen (batch calibration)
Drift at 0% oxygen	< 0.5% O <sub>2</sub> per day (sampling interval of 6 min)
Response time (t <sub>90</sub> )	< 30 s
Temperature range	5 – 50°C
Sensor stability	Sensor material can be degraded by some microorganisms
Cross-sensitivity to	Organic solvents, such as acetone, toluene, chloroform or methylene chloride, Chlorine gas; high concentration of fluorescent molecules in the visible range can interfere (mCherry, tdTomato, dsRed, Nile red); complex media can cause a DO-shift
Basic material	Oxygen sensor RF-230250060+61 (at least stable for 7 days with CertiPUR-buffer) <b>DO sensors are light-sensitive; please protect them from direct light!</b>

**DO calibration**

Calibration	1.0 M Sulfite system (Two-point calibration with oxygen-free environment (sodium sulfite) and air-saturated environment)
Settings	BioLector protocol = LG1-RF-calibration, T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Flower Plate (MTP-48-BOH2)
Calibration device	Hardware ID: BL-09-000F-0032
Calibration phase offset	DO -360.39 (DO Ser. 4302-RD, gain 4)
Date of calibration	2023-03-17

**Sterilization procedure**

Sterilization	Beta irradiation (20 kGy)
Steris Process Run ID	2324-3482
Date of sterilization	2023-03-09

**Contact us**

If you have any questions, contact Beckman Coulter Customer Support Center:

- Worldwide, find out in our website at: [www.beckman.de/support/technical](http://www.beckman.de/support/technical)
- In the USA and Canada, call us at 1-800-369-0333
- Outside the USA and Canada, contact your local Beckman Coulter representative