

**Calibration Data Sheet: pH & DO optodes Lot. No. 1305
(HC filter sets)**

Please enter these **calibration parameters** and the **Lot.-No.** in the BioLection software!

Date of calibration: 2013/03/19

pH calibration parameters

Buffer	150 mM Na-Phosphate buffer, <i>CertiPUR</i> buffer: pH 3.00, pH 4.01, pH 9.00, pH 10.00 (25°C) (20 point calibration)						
Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
φ min	56.59	56.54	56.48	56.42	56.37	56.31	56.25
φ max	14.91	14.92	14.94	14.96	14.97	14.99	15.00
dpH	0.56	0.56	0.56	0.56	0.56	0.56	0.56
pHo	6.32	6.31	6.30	6.29	6.29	6.28	6.27
Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
φ min	56.19	56.14	56.08	56.02	55.96	55.91	55.85
φ max	15.02	15.04	15.05	15.07	15.08	15.10	15.12
dpH	0.56	0.56	0.56	0.55	0.55	0.55	0.55
pHo	6.26	6.26	6.25	6.24	6.23	6.23	6.22
Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
φ min	55.79	55.74	55.68	55.62	55.56	55.51	55.45
φ max	15.13	15.15	15.16	15.18	15.20	15.21	15.23
dpH	0.55	0.55	0.55	0.55	0.55	0.55	0.55
pHo	6.21	6.20	6.19	6.19	6.18	6.17	6.16

Sensor properties

Dynamic range	pH 4.00 - 8.40
Resolution	Up to 0.01 pH (software)
Accuracy	± 0.1 pH at pH 4.25 - 5.00; ± 0.02 pH at pH 5.00 – 7.05; ± 0.2 pH at pH 7.05 - 7.85 (batch calibration)
Response time (t90)	At 25 °C < 30 s
Drift at pH = 7	< 0.005 pH per day (sampling interval of 1 min)
Temperature range	5 °C to 50 °C
Compatibility	Aqueous solutions, ethanol, methanol (max. 10 % v/v)
Cross-sensitivity	Reduced to ionic strength (salinity); a high concentration of fluorescent molecules in the visible range can interfere
Basic material	pH sensor HP8-1248-01 (at least stable for 2 days of cultivation)

calibration

Buffer	<i>CertiPUR</i> Reference Material Buffer solutions Set (pH 3.00 ± 0.01 / pH 4.01 ± 0.015 / pH 9.00 ± 0.01 / pH 10.00 ± 0.03, 20 °C); 150 mM Na-Phosphate buffer (16 solutions)
Settings	BioLector protocol = pH-DO-calibration, T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Flowerplate (MTP-48-BOH)
Calibration device	BioLector CX_1102AB (BL068)
Calibration phase offset	pH 256.3 (pH Ser.3053-hc, gain 21)

**Calibration Data Sheet: pH & DO optodes Lot. No. 1305
(HC filter sets)**

Please enter these **calibration parameters** and the **Lot.-No.** in the BioLecton software!

Date of calibration: 2013/03/19

DO calibration parameters

Buffer	Sulfite system						
Temperature	20°C	21°C	22°C	23°C	24°C	25°C	26°C
ϕ cal0	71.41	71.51	71.61	71.70	71.80	71.90	71.99
ϕ cal100	46.27	46.05	45.82	45.59	45.36	45.13	44.90
Temperature	27°C	28°C	29°C	30°C	31°C	32°C	33°C
ϕ cal0	72.09	72.19	72.28	72.38	72.48	72.57	72.67
ϕ cal100	44.67	44.44	44.22	43.99	43.76	43.53	43.30
Temperature	34°C	35°C	36°C	37°C	38°C	39°C	40°C
ϕ cal0	72.77	72.86	72.96	73.06	73.15	73.25	73.34
ϕ cal100	43.07	42.84	42.61	42.39	42.16	41.93	41.70

Sensor properties

Dynamic range	0 - 100 % air saturation (a.s.)
Resolution	Up to 0.1 % O ₂ (software)
Accuracy	± 2 % dissolved oxygen (batch calibration)
Drift at 0% oxygen	< 0.03% O ₂ within 30 days (sampling interval of 1 min)
Response time (t ₉₀)	< 30 s
Temperature range	0 – 50°C
Cross-sensitivity to	Organic solvents, such as acetone, toluene, chloroform or methylene chloride Chlorine gas
Basic material	Oxygen sensor PSt3-HG-1113-01_3 (at least stable for 2 days of cultivation)

calibration

Calibration	0.5 M Sulfite system (Two-point calibration with oxygen-free environment (nitrogen, sodium sulfite) and air-saturated environment)
Settings	BioLector protocol = pH-DO-calibration, T = 20-40 °C, 800 rpm, 1000 µL/well, shaking diameter 3 mm, MTP-type = Flowerplate (MTP-48-BOH)
Calibration device	BioLector CX_1102AB (BL068)
Calibration phase offset	DO 332.5 (DO Ser.4053-hc, gain 43)

Sterilization procedure

Sterilization	Gamma irradiation (15 kGy)
BGS-certificate No	33103633
Date of sterilization	2013/03/13